

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024130.D
 Acq On : 08 Sep 2021 12:09
 Operator : JC/MD
 Sample : M3675-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-705-707

Quant Time: Sep 09 05:43:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240529	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113862	57.439	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.880%
35) Dibromofluoromethane	5.397	113	87817	55.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.320%
50) Toluene-d8	8.653	98	318813	52.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.240%
62) 4-Bromofluorobenzene	11.085	95	111246	49.873	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.740%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	6755	7.621	ug/l #	81
17) Carbon Disulfide	2.514	76	1325	0.408	ug/l	100
44) Trichloroethene	7.135	130	12292	6.756	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
Data File : VX024130.D
Acq On : 08 Sep 2021 12:09
Operator : JC/MD
Sample : M3675-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-705-707

Quant Time: Sep 09 05:43:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

