

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024132.D
 Acq On : 08 Sep 2021 13:02
 Operator : JC/MD
 Sample : M3675-12DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-720-722DL

Quant Time: Sep 09 05:44:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91186	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108133	54.691	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.380%	
35) Dibromofluoromethane	5.397	113	83609	53.308	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.620%	
50) Toluene-d8	8.653	98	303342	50.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.640%	
62) 4-Bromofluorobenzene	11.085	95	106689	48.549	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.100%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	3854	2.710	ug/l	97
12) 1,1-Dichloroethene	2.325	96	3137	2.307	ug/l #	85
16) Acetone	2.392	43	2367	2.677	ug/l #	86
44) Trichloroethene	7.129	130	86164	48.068	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
Data File : VX024132.D
Acq On : 08 Sep 2021 13:02
Operator : JC/MD
Sample : M3675-12DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-720-722DL

Quant Time: Sep 09 05:44:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

