

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024133.D
 Acq On : 08 Sep 2021 13:25
 Operator : JC/MD
 Sample : M3675-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-TB-20210901

Quant Time: Sep 09 05:44:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	150143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	256758	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	242721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	118393	56.498	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.000%
35) Dibromofluoromethane	5.397	113	92204	54.256	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.520%
50) Toluene-d8	8.653	98	335149	51.816	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.640%
62) 4-Bromofluorobenzene	11.085	95	123964	52.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
Data File : VX024133.D
Acq On : 08 Sep 2021 13:25
Operator : JC/MD
Sample : M3675-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-TB-20210901

Quant Time: Sep 09 05:44:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

