

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031212.D
 Acq On : 08 Sep 2022 15:26
 Operator : JC/MD
 Sample : VX0908WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBL01

Quant Time: Sep 08 18:39:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64170	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	116362	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83677	55.078	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.160%	
35) Dibromofluoromethane	5.391	113	60626	47.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.620%	
50) Toluene-d8	8.647	98	208586	45.276	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.560%	
62) 4-Bromofluorobenzene	11.079	95	77031	45.625	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031212.D
Acq On : 08 Sep 2022 15:26
Operator : JC/MD
Sample : VX0908WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBL01

Quant Time: Sep 08 18:39:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

