

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031216.D
 Acq On : 08 Sep 2022 17:04
 Operator : JC/MD
 Sample : N4381-10DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-20DL

Quant Time: Sep 09 01:10:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63224	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	111152	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	81740	54.608	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.220%	
35) Dibromofluoromethane	5.391	113	55688	45.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 91.960%	
50) Toluene-d8	8.653	98	191324	43.475	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 86.960%	
62) 4-Bromofluorobenzene	11.085	95	74044	45.912	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.820%	
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	3.117	73	89167	23.518	ug/l	# 90
40) Benzene	6.037	78	17897	3.632	ug/l	100
52) Toluene	8.720	92	1132	0.348	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031216.D
Acq On : 08 Sep 2022 17:04
Operator : JC/MD
Sample : N4381-10DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-20DL

Quant Time: Sep 09 01:10:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

