

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031217.D
 Acq On : 08 Sep 2022 17:27
 Operator : JC/MD
 Sample : N4381-12DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-43DL

Quant Time: Sep 09 01:10:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64563	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	119985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	107293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81877	53.565	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.120%
35) Dibromofluoromethane	5.397	113	57212	43.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.520%
50) Toluene-d8	8.653	98	197825	41.643	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	83.280%#
62) 4-Bromofluorobenzene	11.085	95	73408	42.167	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.340%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	109568	28.300	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031217.D
Acq On : 08 Sep 2022 17:27
Operator : JC/MD
Sample : N4381-12DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-43DL

Quant Time: Sep 09 01:10:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

