

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031220.D
 Acq On : 08 Sep 2022 18:37
 Operator : JC/MD
 Sample : N4490-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14-B-SP

Quant Time: Sep 08 19:01:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	61637	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	111388	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87011	59.626	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.260%
35) Dibromofluoromethane	5.385	113	59519	49.034	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	8.653	98	204213	46.306	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.620%
62) 4-Bromofluorobenzene	11.085	95	80214	49.632	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.260%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	37675	74.563	ug/l	# 86
25) 2-Butanone	4.568	43	4603	6.150	ug/l	100
43) Isopropyl Acetate	6.342	43	27940	10.989	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031220.D
Acq On : 08 Sep 2022 18:37
Operator : JC/MD
Sample : N4490-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14-B-SP

Quant Time: Sep 08 19:01:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

