

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031221.D
 Acq On : 08 Sep 2022 19:01
 Operator : JC/MD
 Sample : N4490-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14-B-SP-GR

Quant Time: Sep 08 19:24:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	66099	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	113936	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85523	54.650	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.300%
35) Dibromofluoromethane	5.391	113	59519	47.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.653	98	204274	45.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.560%
62) 4-Bromofluorobenzene	11.079	95	79426	48.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.100%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	20855	38.488	ug/l	# 84
20) Methylene Chloride	2.788	84	5127	4.026	ug/l	96
25) 2-Butanone	4.568	43	3875	4.828	ug/l	90
37) Ethyl Acetate	4.727	43	5795	3.513	ug/l	96
43) Isopropyl Acetate	6.342	43	25633	9.856	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031221.D
Acq On : 08 Sep 2022 19:01
Operator : JC/MD
Sample : N4490-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14-B-SP-GR

Quant Time: Sep 08 19:24:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

