

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031232.D
 Acq On : 08 Sep 2022 23:19
 Operator : JC/MD
 Sample : N4360-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20220824RE

Quant Time: Sep 09 01:12:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67369	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115592	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	81985	51.401	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.800%
35) Dibromofluoromethane	5.391	113	58856	46.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.440%
50) Toluene-d8	8.653	98	193013	42.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.340%#
62) 4-Bromofluorobenzene	11.079	95	76161	45.411	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.820%
Target Compounds						
						Qvalue
44) Trichloroethene	7.129	130	17108	11.877	ug/l	95
64) Tetrachloroethene	9.269	164	1026	0.993	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031232.D
Acq On : 08 Sep 2022 23:19
Operator : JC/MD
Sample : N4360-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20220824RE

Quant Time: Sep 09 01:12:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

