

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031234.D
 Acq On : 09 Sep 2022 00:06
 Operator : JC/MD
 Sample : N4360-10RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D3-20220824RE

Quant Time: Sep 09 01:12:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63181	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109677	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83672	55.936	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.880%
35) Dibromofluoromethane	5.391	113	58076	48.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	8.653	98	193466	44.553	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.100%
62) 4-Bromofluorobenzene	11.079	95	76477	48.058	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.120%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	8333	7.342	ug/l	95
12) 1,1-Dichloroethene	2.313	96	357	0.335	ug/l #	73
27) cis-1,2-Dichloroethene	4.489	96	4431	3.265	ug/l	88
44) Trichloroethene	7.129	130	18919	13.842	ug/l	89
45) 1,2-Dichloropropane	7.434	63	2540	2.126	ug/l #	84
64) Tetrachloroethene	9.275	164	26540	25.381	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031234.D
Acq On : 09 Sep 2022 00:06
Operator : JC/MD
Sample : N4360-10RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D3-20220824RE

Quant Time: Sep 09 01:12:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

