

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031242.D
 Acq On : 09 Sep 2022 03:59
 Operator : JC/MD
 Sample : N4360-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D2-20220823DL

Quant Time: Sep 09 05:36:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	62947	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	109909	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84595	56.764	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.520%
35) Dibromofluoromethane	5.391	113	55816	46.602	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.200%
50) Toluene-d8	8.653	98	187480	43.084	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	86.160%
62) 4-Bromofluorobenzene	11.079	95	74211	46.536	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.080%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	2014	1.781	ug/l	93
12) 1,1-Dichloroethene	2.325	96	1315	1.237	ug/l	# 78
44) Trichloroethene	7.129	130	26476	19.330	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031242.D
Acq On : 09 Sep 2022 03:59
Operator : JC/MD
Sample : N4360-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D2-20220823DL

Quant Time: Sep 09 05:36:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

