

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031240.D
 Acq On : 09 Sep 2022 03:13
 Operator : JC/MD
 Sample : VX0908WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS02

Quant Time: Sep 09 04:08:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa
 Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	36999	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	62574	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	65765	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	59190	67.571	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 135.140%#	
35) Dibromofluoromethane	5.391	113	39707	58.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.460%	
50) Toluene-d8	8.653	98	128757	51.972	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.940%	
62) 4-Bromofluorobenzene	11.079	95	53210	58.607	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 117.220%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	14231	23.934	ug/l	96
3) Chloromethane	1.288	50	8741	19.131	ug/l	98
4) Vinyl Chloride	1.374	62	11927	22.113	ug/l	100
5) Bromomethane	1.605	94	8540	24.782	ug/l	98
6) Chloroethane	1.678	64	8247	27.257	ug/l	92
7) Trichlorofluoromethane	1.886	101	30324	25.802	ug/l	99
8) Diethyl Ether	2.136	74	7960	23.845	ug/l	76
9) 1,1,2-Trichlorotrifluo...	2.325	101	14413	21.686	ug/l	95
10) Methyl Iodide	2.453	142	15010	18.264	ug/l	# 75
11) Tert butyl alcohol	2.989	59	23152	135.501	ug/l	# 84
12) 1,1-Dichloroethene	2.319	96	13441	21.517	ug/l	91
13) Acrolein	2.239	56	12505	102.790	ug/l	98
14) Allyl chloride	2.666	41	16043	19.118	ug/l	# 65
15) Acrylonitrile	3.068	53	33889	101.602	ug/l	94
16) Acetone	2.386	43	35591	117.344	ug/l	88
17) Carbon Disulfide	2.508	76	32496	21.159	ug/l	95
18) Methyl Acetate	2.709	43	16668	21.228	ug/l	# 88
19) Methyl tert-butyl Ether	3.117	73	52946	23.863	ug/l	# 81
20) Methylene Chloride	2.788	84	14713	20.639	ug/l	89
21) trans-1,2-Dichloroethene	3.093	96	14130	20.415	ug/l	99
22) Diisopropyl ether	3.770	45	33614	19.076	ug/l	# 84
23) Vinyl Acetate	3.727	43	153492	103.774	ug/l	# 88
24) 1,1-Dichloroethane	3.611	63	25463	21.785	ug/l	96
25) 2-Butanone	4.568	43	45877	102.116	ug/l	93
26) 2,2-Dichloropropane	4.483	77	18862	18.034	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	16515	20.783	ug/l	97
28) Bromochloromethane	4.910	49	8815	20.799	ug/l	# 66
29) Tetrahydrofuran	5.013	42	26580	100.210	ug/l	# 77
30) Chloroform	5.099	83	31969	23.123	ug/l	98
31) Cyclohexane	5.477	56	18212	18.605	ug/l	# 83
32) 1,1,1-Trichloroethane	5.385	97	31742	25.473	ug/l	97
36) 1,1-Dichloropropene	5.690	75	21375	20.932	ug/l	95
37) Ethyl Acetate	4.721	43	18227	20.120	ug/l	98
38) Carbon Tetrachloride	5.678	117	27790	25.865	ug/l	95
39) Methylcyclohexane	7.379	83	22163	19.003	ug/l	91
40) Benzene	6.044	78	53860	19.415	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
 Data File : VX031240.D
 Acq On : 09 Sep 2022 03:13
 Operator : JC/MD
 Sample : VX0908WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS02

Quant Time: Sep 09 04:08:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa
 Patel

09/09/2022

Supervised By :Mahesh
 Dadoda
 09/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
41) Methacrylonitrile	4.928	41	9562	20.186	ug/l	#	88
42) 1,2-Dichloroethane	6.086	62	29017	25.548	ug/l		92
43) Isopropyl Acetate	6.348	43	28991	20.297	ug/l	#	94
44) Trichloroethene	7.129	130	15480	19.852	ug/l		84
45) 1,2-Dichloropropane	7.427	63	12639	18.539	ug/l		100
46) Dibromomethane	7.580	93	11509	21.289	ug/l		92
47) Bromodichloromethane	7.824	83	25086	23.855	ug/l		99
48) Methyl methacrylate	7.696	41	14455	20.783	ug/l	#	85
49) 1,4-Dioxane	7.683	88	7975	415.350	ug/l		94
51) 4-Methyl-2-Pentanone	8.580	43	86625	99.043	ug/l		92
52) Toluene	8.720	92	36303	19.841	ug/l		97
53) t-1,3-Dichloropropene	8.982	75	23030	21.022	ug/l		100
54) cis-1,3-Dichloropropene	8.366	75	24260	20.627	ug/l		95
55) 1,1,2-Trichloroethane	9.153	97	15183	20.183	ug/l		95
56) Ethyl methacrylate	9.122	69	23606	21.300	ug/l		89
57) 1,3-Dichloropropane	9.311	76	25268	19.440	ug/l		98
58) 2-Chloroethyl Vinyl ether	8.244	63	44634	85.451	ug/l		91
59) 2-Hexanone	9.433	43	68026	102.849	ug/l		98
60) Dibromochloromethane	9.525	129	18832	24.282	ug/l		94
61) 1,2-Dibromoethane	9.610	107	16789	20.084	ug/l		99
64) Tetrachloroethene	9.275	164	14068	20.548	ug/l		96
65) Chlorobenzene	10.079	112	38063	19.658	ug/l		99
66) 1,1,1,2-Tetrachloroethane	10.165	131	16576	21.917	ug/l		94
67) Ethyl Benzene	10.195	91	74531	21.240	ug/l		96
68) m/p-Xylenes	10.305	106	55250	42.476	ug/l		86
69) o-Xylene	10.646	106	26660	20.536	ug/l		78
70) Styrene	10.659	104	44798	21.519	ug/l	#	89
71) Bromoform	10.799	173	12589	24.982	ug/l	#	100
73) Isopropylbenzene	10.963	105	73980	20.254	ug/l		97
74) N-amyl acetate	10.841	43	22166	19.061	ug/l	#	84
75) 1,1,2,2-Tetrachloroethane	11.213	83	21731	19.058	ug/l		99
76) 1,2,3-Trichloropropane	11.244	75	22143m	19.288	ug/l		
77) Bromobenzene	11.201	156	16071	20.244	ug/l		94
78) n-propylbenzene	11.305	91	83140	19.968	ug/l		97
79) 2-Chlorotoluene	11.366	91	53813	20.862	ug/l		92
80) 1,3,5-Trimethylbenzene	11.451	105	64289	20.936	ug/l		95
81) trans-1,4-Dichloro-2-b...	11.018	75	6436	18.080	ug/l	#	81
82) 4-Chlorotoluene	11.457	91	62684	21.003	ug/l		90
83) tert-Butylbenzene	11.713	119	62656	20.698	ug/l		89
84) 1,2,4-Trimethylbenzene	11.750	105	64694	21.671	ug/l		91
85) sec-Butylbenzene	11.890	105	76141	20.309	ug/l		97
86) p-Isopropyltoluene	12.012	119	63324	20.907	ug/l		89
87) 1,3-Dichlorobenzene	11.969	146	29709	20.136	ug/l		96
88) 1,4-Dichlorobenzene	12.042	146	30204	19.906	ug/l		95
89) n-Butylbenzene	12.335	91	54077	20.708	ug/l		99
90) Hexachloroethane	12.542	117	11724	22.528	ug/l		95
91) 1,2-Dichlorobenzene	12.335	146	29123	20.128	ug/l		93
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6994	22.334	ug/l		75
93) 1,2,4-Trichlorobenzene	13.591	180	17470	20.647	ug/l		97
94) Hexachlorobutadiene	13.725	225	7849	21.845	ug/l		97
95) Naphthalene	13.780	128	70157	21.166	ug/l		100
96) 1,2,3-Trichlorobenzene	13.963	180	18497	21.472	ug/l		97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031240.D
Acq On : 09 Sep 2022 03:13
Operator : JC/MD
Sample : VX0908WBS02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Quant Time: Sep 09 04:08:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022

Supervised By :Mahesh
Dadoda
09/09/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:168 Resp: 3699
Ion Ratio Lower Upper
168 100
99 78.8 44.0 66.0

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#2
Dichlorodifluoromethane
Concen: 23.934 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 85 Resp: 14231
Ion Ratio Lower Upper
85 100
87 35.4 16.5 49.5





#3
Chloromethane
Concen: 19.131 ug/l
RT: 1.288 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion: 50 Resp: 874
Ion Ratio Lower Upper
50 100
52 34.0 26.3 39.5


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022



Supervised By :Mahesh
Dadoda
09/09/2022




#4
Vinyl Chloride
Concen: 22.113 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13




Tgt Ion: 62 Resp: 11927
Ion Ratio Lower Upper
62 100
64 31.2 25.2 37.8





#5
Bromomethane
Concen: 24.782 ug/l
RT: 1.605 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 94 Resp: 8540
Ion Ratio Lower Upper
94 100
96 94.2 76.9 115.3

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#6
Chloroethane
Concen: 27.257 ug/l
RT: 1.678 min Scan# 97
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 64 Resp: 8247
Ion Ratio Lower Upper
64 100
66 38.6 27.2 40.8





#7
 Trichlorofluoromethane
 Concen: 25.802 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS02






Tgt Ion:101 Resp: 30324
 Ion Ratio Lower Upper
 101 100
 103 63.1 49.7 74.5

Manual Integrations
 APPROVED




Reviewed By :Krupa
 Patel
 09/09/2022

Supervised By :Mahesh
 Dadoda
 09/09/2022



#8
 Diethyl Ether
 Concen: 23.845 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13


Tgt Ion: 74 Resp: 7960
 Ion Ratio Lower Upper
 74 100
 45 78.5 51.5 154.7





#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.686 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:101 Resp: 1441
Ion Ratio Lower Upper
101 100
85 47.6 34.0 51.0
151 71.3 59.4 89.0

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#10
Methyl Iodide
Concen: 18.264 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:142 Resp: 15010
Ion Ratio Lower Upper
142 100
127 57.7 30.2 45.2#
141 14.7 11.3 16.9





#11
Tert butyl alcohol
Concen: 135.501 ug/l
RT: 2.989 min Scan# 314
Delta R.T. -0.013 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 59 Resp: 2315
Ion Ratio Lower Upper
59 100
57 3.5 7.4 11.2

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#12
1,1-Dichloroethene
Concen: 21.517 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 96 Resp: 13441
Ion Ratio Lower Upper
96 100
61 158.7 117.4 176.0
98 58.9 52.2 78.4





#13
Acrolein
Concen: 102.790 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion: 56 Resp: 1250
Ion Ratio Lower Upper
56 100
55 72.0 56.6 84.8


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022



Supervised By :Mahesh
Dadoda
09/09/2022




#14
Allyl chloride
Concen: 19.118 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13




Tgt Ion: 41 Resp: 16043
Ion Ratio Lower Upper
41 100
39 92.8 48.9 73.3#
76 47.0 26.9 40.3#





#15
Acrylonitrile
Concen: 101.602 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion: 53 Resp: 33889
Ion Ratio Lower Upper
53 100
52 85.2 65.4 98.0
51 41.3 27.7 41.5

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#16
Acetone
Concen: 117.344 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 43 Resp: 35591
Ion Ratio Lower Upper
43 100
58 27.9 27.7 41.5





#17
Carbon Disulfide
Concen: 21.159 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 76 Resp: 32490
Ion Ratio Lower Upper
76 100
78 7.8 7.8 11.6

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#18
Methyl Acetate
Concen: 21.228 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 43 Resp: 16668
Ion Ratio Lower Upper
43 100
74 30.1 19.4 29.2#





#19
Methyl tert-butyl Ether
Concen: 23.863 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion: 73 Resp: 52940
Ion Ratio Lower Upper
73 100
57 14.2 19.0 28.6

Manual Integrations
APPROVED

Reviewed By :Krupa
Patel




09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#20
Methylene Chloride
Concen: 20.639 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 84 Resp: 14713
Ion Ratio Lower Upper
84 100
49 101.2 93.8 140.6
51 34.7 29.4 44.0
86 71.0 51.0 76.4





#21
trans-1,2-Dichloroethene
Concen: 20.415 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 96 Resp: 14130
Ion Ratio Lower Upper
96 100
61 130.3 105.6 158.4
98 63.2 50.1 75.1

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#22
Diisopropyl ether
Concen: 19.076 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.012 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 45 Resp: 33614
Ion Ratio Lower Upper
45 100
43 66.2 65.4 98.0
87 36.6 22.3 33.5#
59 10.8 9.0 13.4





#23
 Vinyl Acetate
 Concen: 103.774 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS02






Tgt Ion: 43 Resp: 153492
 Ion Ratio Lower Upper
 43 100
 86 15.1 8.5 12.7

Manual Integrations
 APPROVED




Reviewed By :Krupa
 Patel
 09/09/2022


Supervised By :Mahesh
 Dadoda
 09/09/2022



#24
 1,1-Dichloroethane
 Concen: 21.785 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13


Tgt Ion: 63 Resp: 25463
 Ion Ratio Lower Upper
 63 100
 98 8.9 3.5 10.5
 100 4.4 2.3 6.9





#25
2-Butanone
Concen: 102.116 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 43 Resp: 4587
Ion Ratio Lower Upper
43 100
72 30.6 21.7 32.5

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#26
2,2-Dichloropropane
Concen: 18.034 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 77 Resp: 18862
Ion Ratio Lower Upper
77 100
97 21.7 13.0 39.0






#27
cis-1,2-Dichloroethene
Concen: 20.783 ug/l
RT: 4.489 min Scan# 558
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02



Tgt Ion: 96 Resp: 1651
Ion Ratio Lower Upper
96 100
61 140.3 0.0 270.0
98 64.5 0.0 127.8

Manual Integrations
APPROVED

Reviewed By :Krupa




Patel

09/09/2022


Supervised By :Mahesh

Dadoda



09/09/2022




#28
Bromochloromethane
Concen: 20.799 ug/l
RT: 4.910 min Scan# 627
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13



Tgt Ion: 49 Resp: 8815
Ion Ratio Lower Upper
49 100
129 1.4 0.0 4.4
130 108.8 62.8 94.2#





#29

Tetrahydrofuran

Concen: 100.210 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX031240.D


Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02



Tgt Ion:	42	Resp:	26580
Ion	Ratio	Lower	Upper
42	100		
72	59.4	36.5	54.7
71	57.8	33.9	50.9

Manual Integrations
APPROVED

Reviewed By :Krupa




Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#30

Chloroform


Concen: 23.123 ug/l

RT: 5.099 min Scan# 658



Delta R.T. 0.006 min


Lab File: VX031240.D

Acq: 09 Sep 2022 03:13




Tgt Ion:	83	Resp:	31969
Ion	Ratio	Lower	Upper
83	100		
85	65.7	51.5	77.3





#31
Cyclohexane
Concen: 18.605 ug/l
RT: 5.477 min Scan# 719
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 56 Resp: 1821
Ion Ratio Lower Upper
56 100
69 33.1 25.4 38.0
84 106.8 68.9 103.3

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#32
1,1,1-Trichloroethane
Concen: 25.473 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 97 Resp: 31742
Ion Ratio Lower Upper
97 100
99 65.6 51.4 77.0
61 41.7 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 67.571 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.006 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02

Tgt Ion: 65 Resp: 59190
 Ion Ratio Lower Upper
 65 100
 67 45.1 0.0 109.6

Manual Integrations
 APPROVED

Reviewed By :Krupa





Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l




RT: 6.763 min Scan# 931


Delta R.T. -0.006 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13


Tgt Ion:114 Resp: 62574
 Ion Ratio Lower Upper
 114 100
 63 18.4 0.0 38.4
 88 17.0 0.0 31.8





#35
Dibromofluoromethane
Concen: 58.231 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:113 Resp: 3970
Ion Ratio Lower Upper
113 100
111 98.5 83.2 124.8
192 15.6 14.8 22.2

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#36
1,1-Dichloropropene
Concen: 20.932 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 75 Resp: 21375
Ion Ratio Lower Upper
75 100
110 34.3 19.3 57.9
77 31.5 24.4 36.6





#37
Ethyl Acetate
Concen: 20.120 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 43 Resp: 1822
Ion Ratio Lower Upper
43 100
61 14.9 11.7 17.5
70 13.2 9.3 13.9


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022



Supervised By :Mahesh
Dadoda
09/09/2022




#38
Carbon Tetrachloride
Concen: 25.865 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13




Tgt Ion:117 Resp: 27790
Ion Ratio Lower Upper
117 100
119 86.2 74.1 111.1
121 32.9 25.8 38.8





#39
Methylcyclohexane
Concen: 19.003 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 83 Resp: 2216
Ion Ratio Lower Upper
83 100
55 61.4 57.7 86.5
98 48.2 36.5 54.7

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#40
Benzene
Concen: 19.415 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 78 Resp: 53860
Ion Ratio Lower Upper
78 100
77 23.4 18.0 27.0





#41
Methacrylonitrile
Concen: 20.186 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 41 Resp: 956
Ion Ratio Lower Upper
41 100
39 58.9 42.9 64.3
67 86.3 60.1 90.1
52 38.1 22.6 34.0#

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#42
1,2-Dichloroethane
Concen: 25.548 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 62 Resp: 29017
Ion Ratio Lower Upper
62 100
98 7.0 0.0 20.2





#43
Isopropyl Acetate
Concen: 20.297 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 43 Resp: 2899
Ion Ratio Lower Upper
43 100
61 22.2 16.5 24.7
87 18.0 11.1 16.7

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#44
Trichloroethene
Concen: 19.852 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion:130 Resp: 15480
Ion Ratio Lower Upper
130 100
95 113.0 0.0 194.8





#45

1,2-Dichloropropane

Concen: 18.539 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02

Tgt Ion: 63 Resp: 12639

Ion Ratio Lower Upper

63 100

65 31.5 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#46

Dibromomethane

Concen: 21.289 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13




Tgt Ion: 93 Resp: 11509


Ion Ratio Lower Upper

93 100

95 80.1 67.4 101.0


174 84.3 75.6 113.4





#47
Bromodichloromethane
Concen: 23.855 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 83 Resp: 25080
Ion Ratio Lower Upper
83 100
85 65.8 53.5 80.3
127 9.0 6.6 10.0

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#48
Methyl methacrylate
Concen: 20.783 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 41 Resp: 14455
Ion Ratio Lower Upper
41 100
69 94.3 68.8 103.2
39 73.5 44.6 66.8





#49

1,4-Dioxane

Concen: 415.350 ug/l

RT: 7.683 min Scan# 1082

Delta R.T. -0.012 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02

Tgt Ion: 88 Resp: 797

Ion	Ratio	Lower	Upper
88	100		
43	31.8	23.8	35.8
58	61.7	53.8	80.6

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#50

Toluene-d8

Concen: 51.972 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. -0.000 min


Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Tgt Ion: 98 Resp: 128757

Ion	Ratio	Lower	Upper
98	100		
100	63.7	52.3	78.5





#51
4-Methyl-2-Pentanone
Concen: 99.043 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13





Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02

Tgt Ion: 43 Resp: 86629
Ion Ratio Lower Upper
43 100
58 44.4 31.4 47.2

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#52
Toluene
Concen: 19.841 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 92 Resp: 36303
Ion Ratio Lower Upper
92 100
91 170.3 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 21.022 ug/l

RT: 8.982 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02

Tgt Ion: 75 Resp: 23030

Ion Ratio Lower Upper

75 100

77 32.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#54

cis-1,3-Dichloropropene

Concen: 20.627 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13




Tgt Ion: 75 Resp: 24260


Ion Ratio Lower Upper

75 100

77 29.8 25.0 37.4


39 51.6 37.5 56.3





#55
1,1,2-Trichloroethane
Concen: 20.183 ug/l
RT: 9.153 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument : MSVOA_X
ClientSampleId : VX0908WBS02






Tgt Ion	Ratio	Lower	Upper
97	100		
83	78.5	68.9	103.3
85	52.3	44.0	66.0
99	60.5	48.3	72.5


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022



Supervised By :Mahesh
Dadoda
09/09/2022




#56
Ethyl methacrylate
Concen: 21.300 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13



Tgt Ion	Ratio	Lower	Upper
69	100		
41	54.6	52.3	78.5
39	39.0	28.2	42.4





#57

1,3-Dichloropropane

Concen: 19.440 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX031240.D

Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02

Tgt Ion: 76 Resp: 2526

Ion Ratio Lower Upper

76 100

78 33.7 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#58

2-Chloroethyl Vinyl ether

Concen: 85.451 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX031240.D




Acq: 09 Sep 2022 03:13


Tgt Ion: 63 Resp: 44634

Ion Ratio Lower Upper

63 100


106 32.5 22.2 33.2





#59
2-Hexanone
Concen: 102.849 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 43 Resp: 68020
Ion Ratio Lower Upper
43 100
58 56.4 27.6 82.8

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#60
Dibromochloromethane
Concen: 24.282 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:129 Resp: 18832
Ion Ratio Lower Upper
129 100
127 72.2 38.5 115.5





#61
1,2-Dibromoethane
Concen: 20.084 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion:107 Resp: 16789
Ion Ratio Lower Upper
107 100
109 94.5 76.6 114.8

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#62
4-Bromofluorobenzene
Concen: 58.607 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 95 Resp: 53210
Ion Ratio Lower Upper
95 100
174 63.7 0.0 149.0
176 62.3 0.0 146.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:117 Resp: 65768
Ion Ratio Lower Upper
117 100
82 64.9 42.7 64.1
119 33.3 25.4 38.2

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#64
Tetrachloroethene
Concen: 20.548 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:164 Resp: 14068
Ion Ratio Lower Upper
164 100
166 111.2 98.2 147.4
129 101.2 80.8 121.2
131 94.4 75.5 113.3





#65
Chlorobenzene
Concen: 19.658 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:112 Resp: 3806
Ion Ratio Lower Upper
112 100
114 30.0 24.6 36.8

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 21.917 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:131 Resp: 16576
Ion Ratio Lower Upper
131 100
133 93.2 46.9 140.8
119 58.8 34.4 103.1





#67
Ethyl Benzene
Concen: 21.240 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 91 Resp: 7453
Ion Ratio Lower Upper
91 100
106 29.8 25.5 38.3

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#68
m/p-Xylenes
Concen: 42.476 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion:106 Resp: 55250
Ion Ratio Lower Upper
106 100
91 219.5 158.2 237.4





#69
o-Xylene
Concen: 20.536 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13





Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02

Tgt Ion:106 Resp: 26660
Ion Ratio Lower Upper
106 100
91 241.8 103.8 311.6

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#70
Styrene
Concen: 21.519 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:104 Resp: 44798
Ion Ratio Lower Upper
104 100
78 61.5 37.8 56.8#
103 53.5 44.5 66.7





#71
Bromoform
Concen: 24.982 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02





Tgt Ion:173 Resp: 12589
Ion Ratio Lower Upper
173 100
175 47.5 23.8 71.2
254 0.0 0.1 0.1

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:152 Resp: 42064
Ion Ratio Lower Upper
152 100
115 84.7 40.5 121.5
150 164.6 0.0 347.0





#73
Isopropylbenzene
Concen: 20.254 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:105 Resp: 73980
Ion Ratio Lower Upper
105 100
120 25.2 13.5 40.4

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#74
N-ethyl acetate
Concen: 19.061 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 43 Resp: 22166
Ion Ratio Lower Upper
43 100
70 53.0 32.3 48.5#
55 28.9 18.4 27.6#
61 29.7 19.2 28.8#





#75
1,1,2,2-Tetrachloroethane
Concen: 19.058 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 83 Resp: 2173
Ion Ratio Lower Upper
83 100
131 11.5 4.5 13.7
85 64.9 32.5 97.5

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#76
1,2,3-Trichloropropane
Concen: 19.288 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion: 75 Resp: 22143
Ion Ratio Lower Upper
75 100
77 41.5 19.8 59.3





#77

Bromobenzene

Concen: 20.244 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. -0.000 min

Lab File: VX031240.D


Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02



Tgt Ion:156 Resp: 1607

Ion Ratio Lower Upper

156 100

77 175.6 82.7 247.9

158 99.3 47.5 142.5

Manual Integrations

APPROVED

Reviewed By :Krupa




Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#78

n-propylbenzene

Concen: 19.968 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX031240.D




Acq: 09 Sep 2022 03:13


Tgt Ion: 91 Resp: 83140

Ion Ratio Lower Upper

91 100


120 21.5 11.6 34.8





#79
2-Chlorotoluene
Concen: 20.862 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 91 Resp: 5381
Ion Ratio Lower Upper
91 100
126 30.1 17.4 52.3

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel


Supervised By :Mahesh
Dadoda
09/09/2022



#80
1,3,5-Trimethylbenzene
Concen: 20.936 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:105 Resp: 64289
Ion Ratio Lower Upper
105 100
120 45.6 24.4 73.2





#81
trans-1,4-Dichloro-2-butene
Concen: 18.080 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 75 Resp: 6430
Ion Ratio Lower Upper
75 100
53 110.5 74.2 111.4
89 56.2 34.8 52.2

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#82
4-Chlorotoluene
Concen: 21.003 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion: 91 Resp: 62684
Ion Ratio Lower Upper
91 100
126 25.1 15.3 45.9





#83
tert-Butylbenzene
Concen: 20.698 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:119 Resp: 62650
Ion Ratio Lower Upper
119 100
91 67.9 28.3 84.9
134 21.6 11.2 33.5

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#84
1,2,4-Trimethylbenzene
Concen: 21.671 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:105 Resp: 64694
Ion Ratio Lower Upper
105 100
120 40.1 22.9 68.8





#85
 sec-Butylbenzene
 Concen: 20.309 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS02






Tgt Ion:105 Resp: 7614
 Ion Ratio Lower Upper
 105 100
 134 18.5 10.0 30.0

Manual Integrations
 APPROVED




Reviewed By :Krupa
 Patel
 09/09/2022


Supervised By :Mahesh
 Dadoda
 09/09/2022



#86
 p-Isopropyltoluene
 Concen: 20.907 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX031240.D
 Acq: 09 Sep 2022 03:13

Tgt Ion:119 Resp: 63324
 Ion Ratio Lower Upper
 119 100
 134 23.2 13.9 41.6
 91 30.1 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 20.136 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX031240.D


Acq: 09 Sep 2022 03:13

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS02



Tgt Ion:	146	Resp:	29709
Ion Ratio	Lower	Upper	
146	100		
111	46.4	20.3	60.8
148	63.2	31.8	95.3

Manual Integrations
APPROVED

Reviewed By :Krupa



Patel

09/09/2022

Supervised By :Mahesh

Dadoda

09/09/2022



#88

1,4-Dichlorobenzene


Concen: 19.906 ug/l

RT: 12.042 min Scan# 1797



Delta R.T. -0.000 min


Lab File: VX031240.D

Acq: 09 Sep 2022 03:13




Tgt Ion:	146	Resp:	30204
Ion Ratio	Lower	Upper	
146	100		
111	48.0	20.3	60.9
148	63.9	32.3	96.8





#89
n-Butylbenzene
Concen: 20.708 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion: 91 Resp: 5407
Ion Ratio Lower Upper
91 100
92 52.1 26.2 78.5
134 23.9 12.6 37.8

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#90
Hexachloroethane
Concen: 22.528 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:117 Resp: 11724
Ion Ratio Lower Upper
117 100
201 64.7 34.2 102.5





#91
1,2-Dichlorobenzene
Concen: 20.128 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:146 Resp: 2912
Ion Ratio Lower Upper
146 100
111 49.1 20.7 62.1
148 61.7 32.0 96.2


Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022



Supervised By :Mahesh
Dadoda
09/09/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 22.334 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13




Tgt Ion: 75 Resp: 6994
Ion Ratio Lower Upper
75 100
155 66.9 45.0 134.8
157 86.2 56.9 170.7





#93
1,2,4-Trichlorobenzene
Concen: 20.647 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:180 Resp: 17470
Ion Ratio Lower Upper
180 100
182 97.7 48.2 144.6
145 37.2 15.9 47.6

Manual Integrations
APPROVED




Reviewed By :Krupa
Patel
09/09/2022


Supervised By :Mahesh
Dadoda
09/09/2022



#94
Hexachlorobutadiene
Concen: 21.845 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13


Tgt Ion:225 Resp: 7849
Ion Ratio Lower Upper
225 100
223 63.0 30.6 91.6
227 60.8 31.9 95.9





#95
Naphthalene
Concen: 21.166 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS02






Tgt Ion:128 Resp: 7015
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.2
129 10.5 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Krupa
Patel
09/09/2022

Supervised By :Mahesh
Dadoda
09/09/2022



#96
1,2,3-Trichlorobenzene
Concen: 21.472 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX031240.D
Acq: 09 Sep 2022 03:13

Tgt Ion:180 Resp: 18497
Ion Ratio Lower Upper
180 100
182 95.5 47.0 141.0
145 37.1 17.0 50.9

