



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090822\
Data File : VX031209.D
Acq On : 08 Sep 2022 13:58
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Title : SW846 8260
Last Update : Wed Sep 07 23:22:41 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	3630	PASS
75	95	30	60	52.4	11054	PASS
95	95	100	100	100.0	21105	PASS
96	95	5	9	6.2	1313	PASS
173	174	0.00	2	0.7	98	PASS
174	95	50	100	62.1	13109	PASS
175	174	5	9	7.2	943	PASS
176	174	95	101	97.3	12755	PASS
177	176	5	9	6.5	830	PASS