

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX090823\
 Data File : VX037601.D
 Acq On : 08 Sep 2023 13:08
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 09 04:03:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 03:56:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115552	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	280816	142.686	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 285.380%#	
35) Dibromofluoromethane	5.391	113	259303	149.277	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 298.560%#	
50) Toluene-d8	8.653	98	945215	148.361	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 296.720%#	
62) 4-Bromofluorobenzene	11.079	95	353671	147.292	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 294.580%#	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	235896	159.423	ug/l	98
3) Chloromethane	1.307	50	246881	149.044	ug/l	99
4) Vinyl Chloride	1.374	62	265218	155.426	ug/l	99
5) Bromomethane	1.593	94	110831	146.576	ug/l	100
6) Chloroethane	1.660	64	76885	84.103	ug/l	96
7) Trichlorofluoromethane	1.867	101	300963	129.775	ug/l	99
8) Diethyl Ether	2.136	74	155663	151.016	ug/l	73
9) 1,1,2-Trichlorotrifluo...	2.319	101	239080	154.132	ug/l	94
10) Methyl Iodide	2.441	142	299993	149.115	ug/l	90
11) Tert butyl alcohol	2.989	59	333390	757.026	ug/l	# 93
12) 1,1-Dichloroethene	2.306	96	241880	154.490	ug/l	# 81
13) Acrolein	2.239	56	332761	768.545	ug/l	97
14) Allyl chloride	2.660	41	324800	147.942	ug/l	# 84
15) Acrylonitrile	3.068	53	715994	745.716	ug/l	98
16) Acetone	2.392	43	553233	683.560	ug/l	# 83
17) Carbon Disulfide	2.501	76	609986	155.559	ug/l	99
18) Methyl Acetate	2.709	43	489625	149.112	ug/l	# 84
19) Methyl tert-butyl Ether	3.117	73	842753	151.823	ug/l	97
20) Methylene Chloride	2.788	84	278440	150.998	ug/l	# 81
21) trans-1,2-Dichloroethene	3.087	96	267295	152.832	ug/l	83
22) Diisopropyl ether	3.770	45	692794	149.017	ug/l	# 73
23) Vinyl Acetate	3.727	43	2573659	775.458	ug/l	# 86
24) 1,1-Dichloroethane	3.605	63	460273	151.924	ug/l	96
25) 2-Butanone	4.568	43	902832	735.530	ug/l	# 81
26) 2,2-Dichloropropane	4.477	77	348712	156.131	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	320203	153.635	ug/l	83
28) Bromochloromethane	4.897	49	192113	145.272	ug/l	# 65
29) Tetrahydrofuran	5.013	42	587959	747.704	ug/l	# 78
30) Chloroform	5.099	83	493994	151.970	ug/l	96
31) Cyclohexane	5.470	56	377790	157.765	ug/l	86
32) 1,1,1-Trichloroethane	5.385	97	449099	160.410	ug/l	95
36) 1,1-Dichloropropene	5.690	75	359853	156.810	ug/l	96
37) Ethyl Acetate	4.721	43	353488	150.234	ug/l	# 90
38) Carbon Tetrachloride	5.678	117	395251	166.652	ug/l	98
39) Methylcyclohexane	7.379	83	460049	164.413	ug/l	89
40) Benzene	6.037	78	1088911	154.216	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090823\
 Data File : VX037601.D
 Acq On : 08 Sep 2023 13:08
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 09 04:03:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 03:56:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	185468	154.735	ug/l #	84
42) 1,2-Dichloroethane	6.086	62	356732	150.877	ug/l	90
43) Isopropyl Acetate	6.348	43	563623	159.932	ug/l #	87
44) Trichloroethene	7.129	130	302969	156.208	ug/l	93
45) 1,2-Dichloropropane	7.434	63	270778	156.590	ug/l	98
46) Dibromomethane	7.580	93	203954	157.157	ug/l	91
47) Bromodichloromethane	7.824	83	389263	164.148	ug/l	97
48) Methyl methacrylate	7.696	41	268755	159.339	ug/l #	76
49) 1,4-Dioxane	7.665	88	163881	3096.672	ug/l #	85
51) 4-Methyl-2-Pentanone	8.580	43	1633331	728.741	ug/l #	84
52) Toluene	8.720	92	708449	154.474	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	429135	174.280	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	462615	168.863	ug/l #	81
55) 1,1,2-Trichloroethane	9.153	97	289836	151.998	ug/l	97
56) Ethyl methacrylate	9.116	69	451063	162.351	ug/l #	78
57) 1,3-Dichloropropane	9.311	76	476066	149.808	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	1094963	778.946	ug/l	90
59) 2-Hexanone	9.433	43	1267652	741.558	ug/l	80
60) Dibromochloromethane	9.525	129	323666	170.477	ug/l	99
61) 1,2-Dibromoethane	9.610	107	316622	155.057	ug/l	100
64) Tetrachloroethene	9.275	164	248553	157.304	ug/l	97
65) Chlorobenzene	10.079	112	761982	153.702	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	292891	166.705	ug/l	99
67) Ethyl Benzene	10.195	91	1331976	158.063	ug/l	98
68) m/p-Xylenes	10.305	106	1040143	313.776	ug/l	94
69) o-Xylene	10.640	106	519119	157.495	ug/l	94
70) Styrene	10.659	104	861355	160.735	ug/l	96
71) Bromoform	10.799	173	244171	150.554	ug/l #	100
73) Isopropylbenzene	10.963	105	1319877	152.067	ug/l	98
74) N-amyl acetate	10.848	43	475977	159.089	ug/l #	84
75) 1,1,2,2-Tetrachloroethane	11.213	83	447060	148.216	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	505993	186.656	ug/l	89
77) Bromobenzene	11.201	156	324682	151.126	ug/l	90
78) n-propylbenzene	11.305	91	1495161	154.898	ug/l	97
79) 2-Chlorotoluene	11.366	91	908304	148.096	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	1124025	152.679	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	149651	149.805	ug/l #	78
82) 4-Chlorotoluene	11.457	91	1032168	149.255	ug/l	96
83) tert-Butylbenzene	11.719	119	1152518	155.278	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	1129188	153.411	ug/l	95
85) sec-Butylbenzene	11.890	105	1391739	155.541	ug/l	98
86) p-Isopropyltoluene	12.012	119	1193329	157.476	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	609503	149.106	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	616924	148.354	ug/l	99
89) n-Butylbenzene	12.335	91	1023754	164.424	ug/l	97
90) Hexachloroethane	12.542	117	222942	173.135	ug/l	89
91) 1,2-Dichlorobenzene	12.335	146	614788	150.984	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	111728	173.978	ug/l	72
93) 1,2,4-Trichlorobenzene	13.591	180	411396	167.042	ug/l	99
94) Hexachlorobutadiene	13.725	225	179370	164.992	ug/l	99
95) Naphthalene	13.774	128	1447333	168.232	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	416424	163.694	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090823\
Data File : VX037601.D
Acq On : 08 Sep 2023 13:08
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Sep 09 04:03:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 03:56:07 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090823\
Data File : VX037601.D
Acq On : 08 Sep 2023 13:08
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Sep 09 04:03:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 03:56:07 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 145797
Ion Ratio Lower Upper
168 100
99 58.8 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 159.423 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 85 Resp: 235896
Ion Ratio Lower Upper
85 100
87 32.1 16.7 50.1





#3

Chloromethane

Concen: 149.044 ug/l

RT: 1.307 min Scan# 30

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 246881

Ion Ratio Lower Upper

50 100

52 32.7 25.5 38.3



#4

Vinyl Chloride

Concen: 155.426 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 62 Resp: 265218

Ion Ratio Lower Upper

62 100

64 32.4 25.4 38.0









#5
Bromomethane
Concen: 146.576 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 110831
Ion Ratio Lower Upper
94 100
96 93.6 74.6 111.8



#6
Chloroethane
Concen: 84.103 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.013 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 64 Resp: 76885
Ion Ratio Lower Upper
64 100
66 33.8 25.1 37.7








#7
Trichlorofluoromethane
Concen: 129.775 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.012 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:101 Resp: 300963
Ion Ratio Lower Upper
101 100
103 65.3 51.4 77.2



#8
Diethyl Ether
Concen: 151.016 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 74 Resp: 155663
Ion Ratio Lower Upper
74 100
45 78.3 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 154.132 ug/l
RT: 2.319 min Scan# 203
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 239080
Ion Ratio Lower Upper
101 100
85 41.9 37.0 55.6
151 77.8 58.6 87.8



#10
Methyl Iodide
Concen: 149.115 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion:142 Resp: 299993
Ion Ratio Lower Upper
142 100
127 43.6 41.6 62.4
141 14.1 11.6 17.4









#11
Tert butyl alcohol
Concen: 757.026 ug/l
RT: 2.989 min Scan# 312
Delta R.T. 0.018 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 333390
Ion Ratio Lower Upper
59 100
57 10.0 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 154.490 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 96 Resp: 241880
Ion Ratio Lower Upper
96 100
61 144.4 144.6 216.8#
98 62.7 50.4 75.6





#13

Acrolein

Concen: 768.545 ug/l

RT: 2.239 min Scan# 189

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 332761

Ion Ratio Lower Upper

56 100

55 69.1 57.1 85.7



#14

Allyl chloride

Concen: 147.942 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion: 41 Resp: 324800


Ion Ratio Lower Upper

41 100

39 60.5 55.8 83.6

76 45.9 24.3 36.5





#15

Acrylonitrile

Concen: 745.716 ug/l

RT: 3.068 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 53 Resp: 715994

Ion Ratio Lower Upper

53 100

52 82.0 66.0 99.0

51 34.8 29.8 44.6



#16

Acetone

Concen: 683.560 ug/l

RT: 2.392 min Scan# 214

Delta R.T. 0.006 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 43 Resp: 553233

Ion Ratio Lower Upper

43 100


58 36.5 21.9 32.9#








#17
Carbon Disulfide
Concen: 155.559 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 609986
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 149.112 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 43 Resp: 489625
Ion Ratio Lower Upper
43 100
74 30.1 17.8 26.8#









#19
Methyl tert-butyl Ether
Concen: 151.823 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 842753
Ion Ratio Lower Upper
73 100
57 20.2 17.3 25.9



#20
Methylene Chloride
Concen: 150.998 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 84 Resp: 278440
Ion Ratio Lower Upper
84 100
49 97.7 103.6 155.4#
51 31.3 33.8 50.6#
86 63.9 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 152.832 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 267295
Ion Ratio Lower Upper
96 100
61 121.3 119.2 178.8
98 64.6 48.7 73.1




#22
Diisopropyl ether
Concen: 149.017 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08



Tgt Ion: 45 Resp: 692794
Ion Ratio Lower Upper
45 100
43 57.7 69.7 104.5#
87 35.7 20.8 31.2#
59 13.5 9.0 13.4#









#23
 Vinyl Acetate
 Concen: 775.458 ug/l
 RT: 3.727 min Scan# 413
 Delta R.T. 0.006 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 2573659
 Ion Ratio Lower Upper
 43 100
 86 14.9 7.8 11.8#



#24
 1,1-Dichloroethane
 Concen: 151.924 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08

Tgt Ion: 63 Resp: 460273
 Ion Ratio Lower Upper
 63 100
 98 8.1 3.3 9.8
 100 5.2 2.1 6.5









#25
2-Butanone
Concen: 735.530 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 902832
Ion Ratio Lower Upper
43 100
72 33.7 19.4 29.2#



#26
2,2-Dichloropropane
Concen: 156.131 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 77 Resp: 348712
Ion Ratio Lower Upper
77 100
97 24.9 10.8 32.3








#27
cis-1,2-Dichloroethene
Concen: 153.635 ug/l
RT: 4.489 min Scan# 511
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 320203
Ion Ratio Lower Upper
96 100
61 125.5 0.0 312.2
98 64.5 0.0 129.0



#28
Bromochloromethane
Concen: 145.272 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 49 Resp: 192113
Ion Ratio Lower Upper
49 100
129 2.4 0.0 3.8
130 101.5 57.5 86.3#








#29
Tetrahydrofuran
Concen: 747.704 ug/l
RT: 5.013 min Scan# 643
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 42 Resp: 587959
Ion Ratio Lower Upper
42 100
72 57.4 34.2 51.4#
71 52.7 32.2 48.2#



#30
Chloroform
Concen: 151.970 ug/l
RT: 5.099 min Scan# 658
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 83 Resp: 493994
Ion Ratio Lower Upper
83 100
85 65.8 50.3 75.5








#31
Cyclohexane
Concen: 157.765 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 377790
Ion Ratio Lower Upper
56 100
69 34.2 23.4 35.2
84 96.3 65.9 98.9



#32
1,1,1-Trichloroethane
Concen: 160.410 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 97 Resp: 449099
Ion Ratio Lower Upper
97 100
99 64.1 51.1 76.7
61 37.3 36.4 54.6









#33
1,2-Dichloroethane-d4
Concen: 142.686 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 280816
Ion Ratio Lower Upper
65 100
67 53.7 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 114 Resp: 248350
Ion Ratio Lower Upper
114 100
63 17.9 0.0 45.0
88 16.6 0.0 36.6









#35
Dibromofluoromethane
Concen: 149.277 ug/l
RT: 5.391 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 259303
Ion Ratio Lower Upper
113 100
111 103.1 82.9 124.3
192 19.7 13.3 19.9



#36
1,1-Dichloropropene
Concen: 156.810 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 75 Resp: 359853
Ion Ratio Lower Upper
75 100
110 38.1 17.0 50.9
77 31.2 24.7 37.1





#37

Ethyl Acetate

Concen: 150.234 ug/l

RT: 4.721 min Scan# 596

Delta R.T. -0.000 min


Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 353488

Ion Ratio Lower Upper

43 100

61 16.7 10.6 16.0#

70 14.8 8.3 12.5#



#38

Carbon Tetrachloride

Concen: 166.652 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion:117 Resp: 395251


Ion Ratio Lower Upper

117 100

119 95.6 78.2 117.2


121 30.8 24.4 36.6








#39
Methylcyclohexane
Concen: 164.413 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 460049
Ion Ratio Lower Upper
83 100
55 68.2 64.6 97.0
98 50.3 36.7 55.1



#40
Benzene
Concen: 154.216 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 78 Resp: 1088911
Ion Ratio Lower Upper
78 100
77 23.8 19.3 28.9








#41
Methacrylonitrile
Concen: 154.735 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 185468
Ion Ratio Lower Upper
41 100
39 54.3 45.2 67.8
67 90.8 52.6 78.8#
52 31.7 22.7 34.1



#42
1,2-Dichloroethane
Concen: 150.877 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 62 Resp: 356732
Ion Ratio Lower Upper
62 100
98 11.5 0.0 16.0





#43

Isopropyl Acetate

Concen: 159.932 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.006 min


Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 563623

Ion Ratio Lower Upper

43 100

61 24.5 15.2 22.8#

87 18.0 9.8 14.6#



#44

Trichloroethene

Concen: 156.208 ug/l

RT: 7.129 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion:130 Resp: 302969

Ion Ratio Lower Upper

130 100

95 95.5 0.0 204.4









#45
1,2-Dichloropropane
Concen: 156.590 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 270778
Ion Ratio Lower Upper
63 100
65 31.7 24.3 36.5



#46
Dibromomethane
Concen: 157.157 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 93 Resp: 203954
Ion Ratio Lower Upper
93 100
95 83.7 66.6 100.0
174 103.2 69.4 104.0









#47
Bromodichloromethane
Concen: 164.148 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 389263
Ion Ratio Lower Upper
83 100
85 64.6 49.8 74.6
127 9.3 7.1 10.7



#48
Methyl methacrylate
Concen: 159.339 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 41 Resp: 268755
Ion Ratio Lower Upper
41 100
69 104.7 59.2 88.8#
39 52.6 47.7 71.5





#49

1,4-Dioxane

Concen: 3096.672 ug/l

RT: 7.665 min Scan# 1079

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 163881

Ion Ratio Lower Upper

88 100

43 24.8 0.0 0.0#

58 62.5 60.1 90.1



#50

Toluene-d8

Concen: 148.361 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 98 Resp: 945215

Ion Ratio Lower Upper

98 100

100 67.3 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 728.741 ug/l

RT: 8.580 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 1633331

Ion Ratio Lower Upper

43 100

58 47.8 30.4 45.6#



#52

Toluene

Concen: 154.474 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 92 Resp: 708449

Ion Ratio Lower Upper

92 100

91 167.0 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 174.280 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. -0.000 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 429135
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 168.863 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08


Tgt Ion: 75 Resp: 462615
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.3 37.9
 39 36.4 46.7 70.1#








#55
1,1,2-Trichloroethane
Concen: 151.998 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Instrument :
MSVOA_X
ClientSampleId :




Tgt Ion: 97 Resp: 289836

Ion	Ratio	Lower	Upper
97	100		
83	84.0	65.8	98.8
85	54.8	41.6	62.4
99	63.7	47.8	71.8






#56
Ethyl methacrylate
Concen: 162.351 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08



Tgt Ion: 69 Resp: 451063

Ion	Ratio	Lower	Upper
69	100		
41	52.3	56.8	85.2#
39	26.4	32.0	48.0#





#57

1,3-Dichloropropane

Concen: 149.808 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. -0.000 min MSVOA_X

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 476066

Ion Ratio Lower Upper

76 100

78 32.8 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 778.946 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 63 Resp: 1094963

Ion Ratio Lower Upper

63 100

106 33.6 22.5 33.7









#59
2-Hexanone
Concen: 741.558 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1267652
Ion Ratio Lower Upper
43 100
58 66.5 26.2 78.5



#60
Dibromochloromethane
Concen: 170.477 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 129 Resp: 323666
Ion Ratio Lower Upper
129 100
127 77.2 38.8 116.4









#61
1,2-Dibromoethane
Concen: 155.057 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 316622
Ion Ratio Lower Upper
107 100
109 94.2 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 147.292 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Tgt Ion: 95 Resp: 353671
Ion Ratio Lower Upper
95 100
174 76.8 0.0 135.6
176 73.8 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 225641
Ion Ratio Lower Upper
117 100
82 55.4 46.2 69.4
119 31.5 25.5 38.3




#64
Tetrachloroethene
Concen: 157.304 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08



Tgt Ion:164 Resp: 248553
Ion Ratio Lower Upper
164 100
166 124.8 102.7 154.1
129 97.3 80.4 120.6
131 93.3 76.5 114.7





#65

Chlorobenzene

Concen: 153.702 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 761982

Ion Ratio Lower Upper

112 100

114 32.3 25.7 38.5



#66

1,1,1,2-Tetrachloroethane

Concen: 166.705 ug/l

RT: 10.165 min Scan# 1489

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion:131 Resp: 292891


Ion Ratio Lower Upper

131 100

133 95.5 47.1 141.4

119 65.0 32.2 96.6









#67
Ethyl Benzene
Concen: 158.063 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 1331976
Ion Ratio Lower Upper
91 100
106 32.3 24.9 37.3



#68
m/p-Xylenes
Concen: 313.776 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 106 Resp: 1040143
Ion Ratio Lower Upper
106 100
91 197.8 165.5 248.3









#69
o-Xylene
Concen: 157.495 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 519119
Ion Ratio Lower Upper
106 100
91 208.4 108.9 326.8



#70
Styrene
Concen: 160.735 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion:104 Resp: 861355
Ion Ratio Lower Upper
104 100
78 48.6 43.6 65.4
103 54.2 43.6 65.4





#71

Bromoform

Concen: 150.554 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 244171

Ion Ratio Lower Upper

173 100

175 48.5 24.1 72.4

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion:152 Resp: 115552


Ion Ratio Lower Upper

152 100

115 136.7 43.3 129.9#

150 208.8 0.0 345.0









#73
Isopropylbenzene
Concen: 152.067 ug/l
RT: 10.963 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 1319877
Ion Ratio Lower Upper
105 100
120 27.0 12.9 38.6



#74
N-ethyl acetate
Concen: 159.089 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 43 Resp: 475977
Ion Ratio Lower Upper
43 100
70 55.4 32.9 49.3#
55 27.6 19.3 28.9
61 30.1 18.8 28.2#





#75

1,1,2,2-Tetrachloroethane

Concen: 148.216 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 447060

Ion Ratio Lower Upper

83 100

131 10.9 5.3 15.8

85 65.1 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 186.656 ug/l

RT: 11.244 min Scan# 1666

Delta R.T. 0.006 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 75 Resp: 505993

Ion Ratio Lower Upper

75 100

77 31.8 19.1 57.5





#77

Bromobenzene

Concen: 151.126 ug/l

RT: 11.201 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 324682

Ion Ratio Lower Upper

156 100

77 155.5 88.4 265.3

158 97.2 48.0 144.0



#78

n-propylbenzene

Concen: 154.898 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 91 Resp: 1495161

Ion Ratio Lower Upper

91 100

120 24.3 11.5 34.4





#79

2-Chlorotoluene

Concen: 148.096 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 908304

Ion Ratio Lower Upper

91 100

126 34.3 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 152.679 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion: 105 Resp: 1124025

Ion Ratio Lower Upper

105 100

120 49.5 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 149.805 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 149651
Ion Ratio Lower Upper
75 100
53 80.7 90.5 135.7#
89 46.9 36.9 55.3



#82
4-Chlorotoluene
Concen: 149.255 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 91 Resp: 1032168
Ion Ratio Lower Upper
91 100
126 30.5 14.2 42.6





#83

tert-Butylbenzene

Concen: 155.278 ug/l

RT: 11.719 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 1152518

Ion Ratio Lower Upper

119 100

91 56.0 31.2 93.6

134 23.0 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 153.411 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VX037601.D




Acq: 08 Sep 2023 13:08


Tgt Ion:105 Resp: 1129188

Ion Ratio Lower Upper

105 100


120 45.6 21.2 63.6








#85
 sec-Butylbenzene
 Concen: 155.541 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:105 Resp: 1391739
 Ion Ratio Lower Upper
 105 100
 134 20.7 9.8 29.4




#86
 p-Isopropyltoluene
 Concen: 157.476 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX037601.D
 Acq: 08 Sep 2023 13:08



Tgt Ion:119 Resp: 1193329
 Ion Ratio Lower Upper
 119 100
 134 26.7 12.8 38.3
 91 23.7 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 149.106 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 609503

Ion Ratio Lower Upper

146 100

111 42.0 22.1 66.1

148 64.5 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 148.354 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion:146 Resp: 616924


Ion Ratio Lower Upper

146 100

111 40.4 21.3 63.7

148 63.5 31.8 95.4





#89

n-Butylbenzene

Concen: 164.424 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08





Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 91 Resp: 1023754

Ion	Ratio	Lower	Upper
91	100		
92	55.9	27.4	82.0
134	27.9	12.6	37.8



#90

Hexachloroethane

Concen: 173.135 ug/l

RT: 12.542 min Scan# 1879




Delta R.T. -0.000 min


Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Tgt Ion: 117 Resp: 222942

Ion	Ratio	Lower	Upper
117	100		
201	74.5	32.9	98.6









#91
1,2-Dichlorobenzene
Concen: 150.984 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 614788
Ion Ratio Lower Upper
146 100
111 43.3 22.4 67.0
148 63.6 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 173.978 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX037601.D
Acq: 08 Sep 2023 13:08

Tgt Ion: 75 Resp: 111728
Ion Ratio Lower Upper
75 100
155 92.7 36.1 108.5
157 118.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 167.042 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 411396

Ion Ratio Lower Upper

180 100

182 96.2 48.3 144.8

145 33.2 18.0 54.0



#94

Hexachlorobutadiene

Concen: 164.992 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08




Tgt Ion:225 Resp: 179370


Ion Ratio Lower Upper

225 100

223 61.1 31.1 93.2

227 63.7 32.5 97.5





#95

Naphthalene

Concen: 168.232 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min


Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:128 Resp: 1447333

Ion Ratio Lower Upper

128 100

127 12.8 10.3 15.5

129 11.2 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 163.694 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX037601.D

Acq: 08 Sep 2023 13:08

Tgt Ion:180 Resp: 416424

Ion Ratio Lower Upper

180 100

182 95.1 46.9 140.6

145 35.0 17.8 53.5

