

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090920\
 Data File : VX018320.D
 Acq On : 09 Sep 2020 14:21
 Operator : JC/SP
 Sample : PB131488TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131488TB

Quant Time: Sep 10 07:14:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	442157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	727868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	719175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	350768	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	318664	51.91	ug/l	0.00
Spiked Amount			Recovery	=	103.82%	
35) Dibromofluoromethane	5.47	113	248724	50.17	ug/l	0.00
Spiked Amount			Recovery	=	100.34%	
50) Toluene-d8	8.70	98	894363	48.44	ug/l	0.00
Spiked Amount			Recovery	=	96.88%	
62) 4-Bromofluorobenzene	11.12	95	346614	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	24474	10.232	ug/l #	86
18) Methyl Acetate	2.75	43	5947	1.030	ug/l	91
20) Methylene Chloride	2.84	84	16737	2.308	ug/l	94
43) Isopropyl Acetate	6.42	43	112753	8.953	ug/l	97
95) Naphthalene	13.82	128	56356	2.430	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090920\
Data File : VX018320.D
Acq On : 09 Sep 2020 14:21
Operator : JC/SP
Sample : PB131488TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131488TB

Quant Time: Sep 10 07:14:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

