

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090920\
 Data File : VX018323.D
 Acq On : 09 Sep 2020 15:29
 Operator : JC/SP
 Sample : PB131488ZHE#12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131488ZHE#12

Quant Time: Sep 10 07:31:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	447823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	717257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	697759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	331343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	320176	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
35) Dibromofluoromethane	5.46	113	247761	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	8.70	98	898448	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
62) 4-Bromofluorobenzene	11.12	95	335152	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
Target Compounds						
16) Acetone	2.42	43	22281	9.198	ug/l	95
20) Methylene Chloride	2.84	84	14251	1.750	ug/l	94
43) Isopropyl Acetate	6.42	43	118597	9.556	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090920\
Data File : VX018323.D
Acq On : 09 Sep 2020 15:29
Operator : JC/SP
Sample : PB131488ZHE#12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131488ZHE#12

Quant Time: Sep 10 07:31:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

