

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090920\
 Data File : VX018327.D
 Acq On : 09 Sep 2020 17:23
 Operator : JC/SP
 Sample : L3902-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-PZ460I-20200901

Quant Time: Sep 10 08:57:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	442093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	677322	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	647269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	307467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	309398	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
35) Dibromofluoromethane	5.47	113	242914	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
50) Toluene-d8	8.70	98	845889	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.12	95	307683	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4353	0.771	ug/l	94
12) 1,1-Dichloroethene	2.36	96	391994	90.732	ug/l	99
24) 1,1-Dichloroethane	3.67	63	305769	35.062	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	54056	9.802	ug/l	87
42) 1,2-Dichloroethane	6.17	62	5306	0.754	ug/l	89
44) Trichloroethene	7.19	130	431330	83.158	ug/l	100
64) Tetrachloroethene	9.32	164	27216	5.722	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090920\
Data File : VX018327.D
Acq On : 09 Sep 2020 17:23
Operator : JC/SP
Sample : L3902-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-PZ460I-20200901

Quant Time: Sep 10 08:57:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

