


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090920\
 Data File : VX018309.D
 Acq On : 09 Sep 2020 09:27
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Title : SW846 8260
 Last Update : Fri Aug 21 03:03:56 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	17020	PASS
75	95	30	60	51.2	49741	PASS
95	95	100	100	100.0	97078	PASS
96	95	5	9	6.3	6098	PASS
173	174	0.00	2	2.0	1565	PASS
174	95	50	100	81.6	79189	PASS
175	174	5	9	7.7	6100	PASS
176	174	95	101	100.7	79730	PASS
177	176	5	9	5.1	4037	PASS