

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090921\
 Data File : VX024144.D
 Acq On : 09 Sep 2021 16:43
 Operator : JC/MD
 Sample : M3694-01 10X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-1

Quant Time: Sep 10 07:29:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229753	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93712	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98078	51.567	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.140%
35) Dibromofluoromethane	5.391	113	74581	49.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	8.653	98	280724	48.503	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.085	95	109220	51.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090921\
Data File : VX024144.D
Acq On : 09 Sep 2021 16:43
Operator : JC/MD
Sample : M3694-01 10X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-1

Quant Time: Sep 10 07:29:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

