

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031247.D
 Acq On : 09 Sep 2022 11:14
 Operator : JC/MD
 Sample : VX0909MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0909MBL01

Quant Time: Sep 09 15:25:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	59237	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	105635	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39572	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88759	63.288	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	126.580%#
35) Dibromofluoromethane	5.385	113	60418	52.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.980%
50) Toluene-d8	8.647	98	201304	48.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.260%
62) 4-Bromofluorobenzene	11.079	95	81276	53.028	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031247.D
Acq On : 09 Sep 2022 11:14
Operator : JC/MD
Sample : VX0909MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0909MBL01

Quant Time: Sep 09 15:25:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

