

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031248.D
 Acq On : 09 Sep 2022 11:42
 Operator : JC/MD
 Sample : VX0909WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0909WBL01

Quant Time: Sep 09 15:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57370	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	100031	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79810	58.759	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.520%	
35) Dibromofluoromethane	5.385	113	52526	48.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.380%	
50) Toluene-d8	8.647	98	182478	46.075	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.160%	
62) 4-Bromofluorobenzene	11.079	95	71463	49.238	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031248.D
Acq On : 09 Sep 2022 11:42
Operator : JC/MD
Sample : VX0909WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0909WBL01

Quant Time: Sep 09 15:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

