

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031250.D
 Acq On : 09 Sep 2022 12:36
 Operator : JC/MD
 Sample : N4505-04DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLI-107DL

Quant Time: Sep 09 15:26:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	58124	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	104320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	93350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90276	65.602	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	131.200%#
35) Dibromofluoromethane	5.385	113	57963	50.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.980%
50) Toluene-d8	8.647	98	194770	47.157	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.320%
62) 4-Bromofluorobenzene	11.085	95	80300	53.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.100%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	31237	65.558	ug/l	# 85
25) 2-Butanone	4.568	43	5100	7.226	ug/l	98
52) Toluene	8.714	92	1091	0.358	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031250.D
Acq On : 09 Sep 2022 12:36
Operator : JC/MD
Sample : N4505-04DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLI-107DL

Quant Time: Sep 09 15:26:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

