

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031252.D
 Acq On : 09 Sep 2022 13:23
 Operator : JC/MD
 Sample : N4505-06DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLI-109DL

Quant Time: Sep 09 15:26:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	56548	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	99061	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	84567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81034	60.527	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.060%
35) Dibromofluoromethane	5.391	113	53021	49.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	8.653	98	175756	44.812	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.620%
62) 4-Bromofluorobenzene	11.079	95	67344	46.854	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.700%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	146579	316.204	ug/l	# 86
25) 2-Butanone	4.562	43	5663	8.247	ug/l	91
52) Toluene	8.720	92	32089	11.078	ug/l	98
67) Ethyl Benzene	10.195	91	1741	0.386	ug/l	# 71
68) m/p-Xylenes	10.299	106	2779	1.661	ug/l	78
69) o-Xylene	10.640	106	1171	0.701	ug/l	79
84) 1,2,4-Trimethylbenzene	11.750	105	594	0.219	ug/l	91
95) Naphthalene	13.774	128	2031	0.673	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031252.D
Acq On : 09 Sep 2022 13:23
Operator : JC/MD
Sample : N4505-06DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLI-109DL

Quant Time: Sep 09 15:26:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

