

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031254.D
 Acq On : 09 Sep 2022 14:10
 Operator : JC/MD
 Sample : N4505-03DL 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CLI-79DL

Quant Time: Sep 09 15:27:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	53717	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	90385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	73400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	33517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74814	58.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.660%
35) Dibromofluoromethane	5.391	113	49788	50.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.100%
50) Toluene-d8	8.647	98	159932	44.692	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.380%
62) 4-Bromofluorobenzene	11.079	95	64961	49.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%
Target Compounds						
16) Acetone	2.386	43	24626	55.924	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031254.D
Acq On : 09 Sep 2022 14:10
Operator : JC/MD
Sample : N4505-03DL 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CLI-79DL

Quant Time: Sep 09 15:27:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

