

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031257.D
 Acq On : 09 Sep 2022 15:21
 Operator : JC/MD
 Sample : N4360-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D-20220824DL

Quant Time: Sep 12 00:51:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	46701	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	81153	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	79601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35877	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70685	63.930	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 127.860%#	
35) Dibromofluoromethane	5.385	113	45940	51.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.900%	
50) Toluene-d8	8.653	98	148769	46.302	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.600%	
62) 4-Bromofluorobenzene	11.079	95	64178	54.505	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.000%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1335	1.591	ug/l	94
12) 1,1-Dichloroethene	2.325	96	233	0.296	ug/l #	60
44) Trichloroethene	7.135	130	14519	14.357	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031257.D
Acq On : 09 Sep 2022 15:21
Operator : JC/MD
Sample : N4360-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D-20220824DL

Quant Time: Sep 12 00:51:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

