

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090922\  
 Data File : VX031265.D  
 Acq On : 09 Sep 2022 18:29  
 Operator : JC/MD  
 Sample : N4433-10  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TB

Quant Time: Sep 12 01:00:23 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X090722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 07 23:22:41 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.556  | 168  | 44686    | 50.000 | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene    | 6.763  | 114  | 76958    | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 66598    | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 28463    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 60092    | 56.800   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 113.600% |
| 35) Dibromofluoromethane    | 5.385  | 113   | 36435    | 43.446   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 86.900%  |
| 50) Toluene-d8              | 8.653  | 98    | 118200   | 38.793   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 77.580%# |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 44280    | 39.656   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 79.320%# |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090922\  
Data File : VX031265.D  
Acq On : 09 Sep 2022 18:29  
Operator : JC/MD  
Sample : N4433-10  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TB

Quant Time: Sep 12 01:00:23 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X090722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 07 23:22:41 2022  
Response via : Initial Calibration

