

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090922\  
 Data File : VX031266.D  
 Acq On : 09 Sep 2022 18:52  
 Operator : JC/MD  
 Sample : N4360-01DL 20X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TT180D2-20220823DL

Quant Time: Sep 12 01:00:40 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X090722W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 07 23:22:41 2022  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|------------------------------|--------|-------|----------|----------|-------|-----------|
| -----                        |        |       |          |          |       |           |
| Internal Standards           |        |       |          |          |       |           |
| 1) Pentafluorobenzene        | 5.556  | 168   | 43656    | 50.000   | ug/l  | # 0.00    |
| 34) 1,4-Difluorobenzene      | 6.769  | 114   | 78575    | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 70820    | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 30246    | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds  |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4    | 5.958  | 65    | 69299    | 67.048   | ug/l  | 0.00      |
| Spiked Amount                | 50.000 | Range | 74 - 125 | Recovery | =     | 134.100%# |
| 35) Dibromofluoromethane     | 5.385  | 113   | 40853    | 47.711   | ug/l  | 0.00      |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =     | 95.420%   |
| 50) Toluene-d8               | 8.653  | 98    | 138341   | 44.469   | ug/l  | 0.00      |
| Spiked Amount                | 50.000 | Range | 86 - 113 | Recovery | =     | 88.940%   |
| 62) 4-Bromofluorobenzene     | 11.085 | 95    | 55116    | 48.344   | ug/l  | 0.00      |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =     | 96.680%   |
| Target Compounds             |        |       |          |          |       |           |
|                              |        |       |          |          |       | Qvalue    |
| 9) 1,1,2-Trichlorotrifluo... | 2.325  | 101   | 1399     | 1.784    | ug/l  | 89        |
| 12) 1,1-Dichloroethene       | 2.319  | 96    | 939      | 1.274    | ug/l  | # 81      |
| 44) Trichloroethene          | 7.129  | 130   | 20274    | 20.705   | ug/l  | 95        |
| -----                        |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090922\  
Data File : VX031266.D  
Acq On : 09 Sep 2022 18:52  
Operator : JC/MD  
Sample : N4360-01DL 20X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TT180D2-20220823DL

Quant Time: Sep 12 01:00:40 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X090722W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 07 23:22:41 2022  
Response via : Initial Calibration

