

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031267.D
 Acq On : 09 Sep 2022 19:16
 Operator : JC/MD
 Sample : N4360-14RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20220823RE

Quant Time: Sep 12 01:00:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	45263	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	76629	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	69700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	29246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72081	67.263	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	134.520%#
35) Dibromofluoromethane	5.391	113	43753	52.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.800%
50) Toluene-d8	8.653	98	139415	45.952	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.900%
62) 4-Bromofluorobenzene	11.079	95	55058	49.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.040%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	854	2.302	ug/l	# 59
20) Methylene Chloride	2.782	84	812	0.931	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031267.D
Acq On : 09 Sep 2022 19:16
Operator : JC/MD
Sample : N4360-14RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20220823RE

Quant Time: Sep 12 01:00:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

