

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
 Data File : VX031269.D
 Acq On : 09 Sep 2022 20:03
 Operator : JC/MD
 Sample : N4360-09DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20220824DL

Quant Time: Sep 12 01:01:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	43609	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	75369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	67518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	29310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68273	66.126	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 132.260%#	
35) Dibromofluoromethane	5.385	113	41602	50.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.300%	
50) Toluene-d8	8.653	98	136841	45.858	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 91.720%	
62) 4-Bromofluorobenzene	11.085	95	53316	48.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.500%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.324	101	620	0.791	ug/l	# 94
44) Trichloroethene	7.128	130	11116	11.835	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090922\
Data File : VX031269.D
Acq On : 09 Sep 2022 20:03
Operator : JC/MD
Sample : N4360-09DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20220824DL

Quant Time: Sep 12 01:01:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

