



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX090922\  
Data File : VX031245.D  
Acq On : 09 Sep 2022 10:10  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X090722W.M  
Title : SW846 8260  
Last Update : Wed Sep 07 23:22:41 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 18.9  | 3244  | PASS      |
| 75     | 95      | 30     | 60     | 57.0  | 9798  | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 17194 | PASS      |
| 96     | 95      | 5      | 9      | 6.8   | 1172  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.7   | 78    | PASS      |
| 174    | 95      | 50     | 100    | 64.7  | 11119 | PASS      |
| 175    | 174     | 5      | 9      | 6.8   | 755   | PASS      |
| 176    | 174     | 95     | 101    | 97.5  | 10841 | PASS      |
| 177    | 176     | 5      | 9      | 6.4   | 696   | PASS      |