

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
 Data File : VX037618.D
 Acq On : 09 Sep 2023 11:18
 Operator : JC/MD
 Sample : VX0909WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0909WBL01

Quant Time: Sep 11 01:18:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 06:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238352	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99784	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82625	43.005	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.020%
35) Dibromofluoromethane	5.385	113	75156	45.081	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.160%
50) Toluene-d8	8.647	98	274117	44.830	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.660%#
62) 4-Bromofluorobenzene	11.079	95	102435	44.450	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
Data File : VX037618.D
Acq On : 09 Sep 2023 11:18
Operator : JC/MD
Sample : VX0909WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0909WBL01

Quant Time: Sep 11 01:18:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 06:38:11 2023
Response via : Initial Calibration

