

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
 Data File : VX037623.D
 Acq On : 09 Sep 2023 13:22
 Operator : JC/MD
 Sample : 04358-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 11 01:21:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 06:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	142428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236460	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79477	41.338	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.680%
35) Dibromofluoromethane	5.385	113	71687	43.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.680%
50) Toluene-d8	8.647	98	276694	45.614	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.220%#
62) 4-Bromofluorobenzene	11.079	95	98787	43.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
Data File : VX037623.D
Acq On : 09 Sep 2023 13:22
Operator : JC/MD
Sample : 04358-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 11 01:21:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 06:38:11 2023
Response via : Initial Calibration

