

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
 Data File : VX037627.D
 Acq On : 09 Sep 2023 14:59
 Operator : JC/MD
 Sample : 04315-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOWER-1

Quant Time: Sep 11 01:22:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 06:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221533	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75951	42.226	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.460%
35) Dibromofluoromethane	5.385	113	67782	43.745	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.480%
50) Toluene-d8	8.647	98	258811	45.540	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.080%#
62) 4-Bromofluorobenzene	11.079	95	89697	41.878	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11444	15.471	ug/l #	87
30) Chloroform	5.099	83	2395	0.806	ug/l	93
47) Bromodichloromethane	7.824	83	1598	0.755	ug/l	94
60) Dibromochloromethane	9.525	129	2467	1.457	ug/l	97
71) Bromoform	10.799	173	1755	1.703	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
Data File : VX037627.D
Acq On : 09 Sep 2023 14:59
Operator : JC/MD
Sample : 04315-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOWER-1

Quant Time: Sep 11 01:22:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 06:38:11 2023
Response via : Initial Calibration

