

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
 Data File : VX037629.D
 Acq On : 09 Sep 2023 15:48
 Operator : JC/MD
 Sample : 04296-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NGTW1

Quant Time: Sep 11 01:23:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 09 06:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216022	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80781	41.670	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.340%
35) Dibromofluoromethane	5.385	113	72447	43.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.020%
50) Toluene-d8	8.647	98	282181	46.207	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.420%
62) 4-Bromofluorobenzene	11.079	95	99889	43.400	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.800%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2579	3.235	ug/l #	70
30) Chloroform	5.098	83	2696	0.842	ug/l	93
67) Ethyl Benzene	10.201	91	2166	0.268	ug/l	94
78) n-propylbenzene	11.305	91	1780	0.227	ug/l	98
95) Naphthalene	13.780	128	2220	0.318	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
Data File : VX037629.D
Acq On : 09 Sep 2023 15:48
Operator : JC/MD
Sample : 04296-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NGTW1

Quant Time: Sep 11 01:23:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 09 06:38:11 2023
Response via : Initial Calibration

