



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090923\
Data File : VX037615.D
Acq On : 09 Sep 2023 09:53
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090823W.M
Title : SW846 8260
Last Update : Mon Sep 11 18:51:41 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	3228	PASS
75	95	30	60	47.7	9989	PASS
95	95	100	100	100.0	20933	PASS
96	95	5	9	6.8	1431	PASS
173	174	0.00	2	0.5	87	PASS
174	95	50	100	75.8	15869	PASS
175	174	5	9	7.1	1133	PASS
176	174	95	101	95.5	15147	PASS
177	176	5	9	6.9	1040	PASS