

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091018\
 Data File : VX004443.D
 Acq On : 10 Sep 2018 12:48
 Operator : JC/MD
 Sample : PB112704TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB112704TB

Quant Time: Sep 11 01:38:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	342949	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	565034	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	566041	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	290298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	240071	53.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.42%	
35) Dibromofluoromethane	5.49	113	212445	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	8.71	98	811095	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.14	95	268662	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
Target Compounds						
16) Acetone	2.44	43	25173	5.91	ug/l	99
18) Methyl Acetate	2.77	43	7560	1.14	ug/l	97
20) Methylene Chloride	2.85	84	17802	4.14	ug/l	95
43) Isopropyl Acetate	6.46	43	224381	23.37	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091018\
Data File : VX004443.D
Acq On : 10 Sep 2018 12:48
Operator : JC/MD
Sample : PB112704TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB112704TB

Quant Time: Sep 11 01:38:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

