

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091019\
 Data File : VX012289.D
 Acq On : 11 Sep 2019 07:48
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 11 08:11:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.409	0.376	8.1	78	0.00
3 P	Chloromethane	0.368	0.381	-3.5	85	0.00
4 C	Vinyl Chloride	0.365	0.353	3.3#	81	0.00
5 T	Bromomethane	0.179	0.196	-9.5	97	0.00
6 T	Chloroethane	0.244	0.246	-0.8	83	0.00
7 T	Trichlorofluoromethane	0.661	0.616	6.8	76	0.00
8 T	Diethyl Ether	0.226	0.237	-4.9	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.446	0.439	1.6	75	0.00
10 T	Methyl Iodide	0.424	0.280	34.0#	49#	0.00
11 T	Tert butyl alcohol	0.079	0.086	-8.9	91	0.00
12 CM	1,1-Dichloroethene	0.378	0.388	-2.6#	80	0.00
13 T	Acrolein	0.037	0.021	43.2#	45#	0.00
14 T	Allyl chloride	0.771	0.759	1.6	76	0.00
15 T	Acrylonitrile	0.139	0.157	-12.9	86	0.00
16 T	Acetone	0.167	0.139	16.8	68	0.00
17 T	Carbon Disulfide	0.794	0.781	1.6	78	0.00
18 T	Methyl Acetate	0.352	0.400	-13.6	89	0.00
19 T	Methyl tert-butyl Ether	1.494	1.603	-7.3	80	0.00
20 T	Methylene Chloride	0.576	0.553	4.0	83	0.00
21 T	trans-1,2-Dichloroethene	0.451	0.464	-2.9	80	0.00
22 T	Diisopropyl ether	1.764	1.861	-5.5	80	0.00
23 T	Vinyl Acetate	1.082	1.085	-0.3	74	0.00
24 P	1,1-Dichloroethane	0.945	0.991	-4.9	81	0.00
25 T	2-Butanone	0.214	0.219	-2.3	81	0.00
26 T	2,2-Dichloropropane	0.887	0.669	24.6#	58	0.00
27 T	cis-1,2-Dichloroethene	0.583	0.615	-5.5	82	0.00
28 T	Bromochloromethane	0.443	0.479	-8.1	80	0.00
29 T	Tetrahydrofuran	0.115	0.135	-17.4	87	0.00
30 C	Chloroform	1.066	1.126	-5.6#	82	0.00
31 T	Cyclohexane	0.656	0.616	6.1	73	0.00
32 T	1,1,1-Trichloroethane	0.915	0.937	-2.4	80	0.00
33 S	1,2-Dichloroethane-d4	0.707	0.742	-5.0	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.321	0.308	4.0	81	0.00
36 T	1,1-Dichloropropene	0.365	0.356	2.5	79	0.00
37 T	Ethyl Acetate	0.225	0.244	-8.4	84	0.00
38 T	Carbon Tetrachloride	0.428	0.399	6.8	76	0.00
39 T	Methylcyclohexane	0.400	0.357	10.8	73	0.00
40 TM	Benzene	1.103	1.099	0.4	81	0.00
41 T	Methacrylonitrile	0.139	0.155	-11.5	88	0.00
42 TM	1,2-Dichloroethane	0.436	0.441	-1.1	81	0.00
43 T	Isopropyl Acetate	0.475	0.498	-4.8	82	0.00
44 TM	Trichloroethene	0.299	0.300	-0.3	83	0.00
45 C	1,2-Dichloropropane	0.305	0.304	0.3#	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VO091019\
 Data File : VX012289.D
 Acq On : 11 Sep 2019 07:48
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 11 08:11:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.167	0.178	-6.6	84	0.00
47 T	Bromodichloromethane	0.456	0.460	-0.9	79	0.00
48 T	Methyl methacrylate	0.227	0.242	-6.6	83	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	98	0.00
50 S	Toluene-d8	1.124	1.141	-1.5	81	0.00
51 T	4-Methyl-2-Pentanone	0.251	0.264	-5.2	84	0.00
52 CM	Toluene	0.729	0.715	1.9#	79	0.00
53 T	t-1,3-Dichloropropene	0.455	0.414	9.0	72	0.00
54 T	cis-1,3-Dichloropropene	0.504	0.459	8.9	73	0.00
55 T	1,1,2-Trichloroethane	0.259	0.265	-2.3	82	0.00
56 T	Ethyl methacrylate	0.350	0.362	-3.4	81	0.00
57 T	1,3-Dichloropropane	0.442	0.455	-2.9	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.179	0.181	-1.1	78	0.00
59 T	2-Hexanone	0.174	0.183	-5.2	81	0.00
60 T	Dibromochloromethane	0.302	0.304	-0.7	77	0.00
61 T	1,2-Dibromoethane	0.241	0.249	-3.3	84	0.00
62 S	4-Bromofluorobenzene	0.521	0.480	7.9	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.264	0.321	-21.6#	102	0.00
65 PM	Chlorobenzene	0.914	0.887	3.0	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.344	0.342	0.6	79	0.00
67 C	Ethyl Benzene	1.625	1.591	2.1#	81	0.00
68 T	m/p-Xylenes	0.598	0.584	2.3	80	0.00
69 T	o-Xylene	0.586	0.572	2.4	80	0.00
70 T	Styrene	1.039	1.035	0.4	81	0.00
71 P	Bromoform	0.179	0.178	0.6	77	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.292	3.175	3.6	79	0.00
74 T	N-amyl acetate	0.979	0.868	11.3	70	0.00
75 P	1,1,2,2-Tetrachloroethane	0.707	0.737	-4.2	83	0.00
76 T	1,2,3-Trichloropropane	0.642	0.647	-0.8	80	0.00
77 T	Bromobenzene	0.758	0.759	-0.1	81	0.00
78 T	n-propylbenzene	3.975	3.781	4.9	76	0.00
79 T	2-Chlorotoluene	2.463	2.373	3.7	79	0.00
80 T	1,3,5-Trimethylbenzene	2.960	2.851	3.7	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.175	18.2	64	0.00
82 T	4-Chlorotoluene	2.960	2.893	2.3	79	0.00
83 T	tert-Butylbenzene	2.874	2.622	8.8	73	0.00
84 T	1,2,4-Trimethylbenzene	3.060	2.914	4.8	78	0.00
85 T	sec-Butylbenzene	3.430	3.234	5.7	77	0.00
86 T	p-Isopropyltoluene	3.169	2.777	12.4	71	0.00
87 T	1,3-Dichlorobenzene	1.544	1.512	2.1	80	0.00
88 T	1,4-Dichlorobenzene	1.561	1.516	2.9	81	0.00
89 T	n-Butylbenzene	3.106	2.923	5.9	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091019\
Data File : VX012289.D
Acq On : 11 Sep 2019 07:48
Operator : SY/SP
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA X/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 11 08:11:42 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.584	0.567	2.9	77	0.00
91 T	1,2-Dichlorobenzene	1.428	1.429	-0.1	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.144	0.146	-1.4	82	0.00
93 T	1,2,4-Trichlorobenzene	0.986	0.933	5.4	75	0.00
94 T	Hexachlorobutadiene	0.491	0.486	1.0	81	0.00
95 T	Naphthalene	2.116	2.204	-4.2	83	0.00
96 T	1,2,3-Trichlorobenzene	0.873	0.863	1.1	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6