

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091020\
 Data File : VX018339.D
 Acq On : 10 Sep 2020 15:15
 Operator : JC/SP
 Sample : L3952-01 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40557

Quant Time: Sep 11 07:16:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	479011	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	760000	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	691255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	341773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	345750	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
35) Dibromofluoromethane	5.46	113	270829	52.32	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	104.64%	
50) Toluene-d8	8.70	98	918083	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
62) 4-Bromofluorobenzene	11.12	95	333884	44.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.04%	
Target Compounds						
16) Acetone	2.42	43	968435	373.739	ug/l	96
25) 2-Butanone	4.63	43	15842	3.720	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091020\
Data File : VX018339.D
Acq On : 10 Sep 2020 15:15
Operator : JC/SP
Sample : L3952-01 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40557

Quant Time: Sep 11 07:16:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

