

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091020\
 Data File : VX018343.D
 Acq On : 10 Sep 2020 16:48
 Operator : JC/SP
 Sample : L3952-02 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40614

Quant Time: Sep 11 07:28:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	419362	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	664574	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	649048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	338640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	294710	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
35) Dibromofluoromethane	5.46	113	226980	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	8.70	98	795799	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
62) 4-Bromofluorobenzene	11.12	95	324204	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
Target Compounds						
16) Acetone	2.42	43	112083	49.408	ug/l	99
40) Benzene	6.11	78	8202	0.455	ug/l	92
52) Toluene	8.76	92	8889	0.778	ug/l	98
86) p-Isopropyltoluene	12.05	119	5872	0.310	ug/l #	68
95) Naphthalene	13.82	128	6839	0.305	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091020\
Data File : VX018343.D
Acq On : 10 Sep 2020 16:48
Operator : JC/SP
Sample : L3952-02 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40614

Quant Time: Sep 11 07:28:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

