

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091020\
 Data File : VX018345.D
 Acq On : 10 Sep 2020 17:34
 Operator : JC/SP
 Sample : L3953-01 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N2930S

Quant Time: Sep 11 07:34:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	440181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	697409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	667057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	353903	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	304122	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
35) Dibromofluoromethane	5.47	113	235064	49.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.98%	
50) Toluene-d8	8.70	98	836439	47.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.56%	
62) 4-Bromofluorobenzene	11.12	95	350044	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	

Target Compounds

					Qvalue	
16) Acetone	2.42	43	5220	2.192	ug/l	# 81
27) cis-1,2-Dichloroethene	4.56	96	4208	0.766	ug/l	95
40) Benzene	6.11	78	1984620	104.999	ug/l	98
52) Toluene	8.77	92	59870	4.995	ug/l	98
67) Ethyl Benzene	10.24	91	914669	40.371	ug/l	99
68) m/p-Xylenes	10.34	106	186212	22.020	ug/l	95
69) o-Xylene	10.68	106	94160	11.486	ug/l	91
73) Isopropylbenzene	11.00	105	53271	2.282	ug/l	99
78) n-propylbenzene	11.34	91	38149	1.439	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	103869	5.415	ug/l	97
84) 1,2,4-Trimethylbenzene	11.79	105	342657	17.471	ug/l	99
86) p-Isopropyltoluene	12.05	119	6622	0.335	ug/l	# 70
95) Naphthalene	13.82	128	8087566	345.571	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091020\
Data File : VX018345.D
Acq On : 10 Sep 2020 17:34
Operator : JC/SP
Sample : L3953-01 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N2930S

Quant Time: Sep 11 07:34:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

