

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091020\
 Data File : VX018335.D
 Acq On : 10 Sep 2020 11:42
 Operator : JC/SP
 Sample : VX0910WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBL01

Quant Time: Sep 11 07:01:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	423487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	689233	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	646485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	298627	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	314359	53.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.94%	
35) Dibromofluoromethane	5.46	113	248909	53.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.06%	
50) Toluene-d8	8.70	98	835303	47.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.56%	
62) 4-Bromofluorobenzene	11.12	95	297415	43.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.48%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091020\
Data File : VX018335.D
Acq On : 10 Sep 2020 11:42
Operator : JC/SP
Sample : VX0910WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0910WBL01

Quant Time: Sep 11 07:01:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX018335.D

Acq: 10 Sep 2020 11:42

Instrument :

MSVOA_X

ClientSampleId :


VX0910WBL01

Tot Ion:168 Resp: 423487

Ion Ratio Lower Upper

168 100

99 59.5 46.3 69.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 53.471 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX018335.D


Acq: 10 Sep 2020 11:42

Tgt Ion: 65 Resp: 314359

Ion Ratio Lower Upper

65 100

67 51.1 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0


6.03


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.01 min

Lab File: VX018335.D

Acq: 10 Sep 2020 11:42

Instrument :

MSVOA_X

ClientSampleId :

VX0910WBL01


Tot Ion:114 Resp: 689233

Ion Ratio Lower Upper

114 100

63 20.3 0.0 42.6




88 15.7 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 53.027 ug/l

RT: 5.46 min Scan# 718

Delta R.T. -0.01 min

Lab File: VX018335.D

Acq: 10 Sep 2020 11:42


Tgt Ion:113 Resp: 248909

Ion Ratio Lower Upper

113 100

111 101.8 81.6 122.4



192 19.4 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.778 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX018335.D

Acq: 10 Sep 2020 11:42

Instrument :

MSVOA_X

ClientSampleId :


VX0910WBL01

Tot Ion: 98 Resp: 835303

Ion Ratio Lower Upper

98 100

100 63.8 52.2 78.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.70

600000

500000

400000



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 43.238 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX018335.D

Acq: 10 Sep 2020 11:42


Tgt Ion: 95 Resp: 297415

Ion Ratio Lower Upper

95 100

174 80.5 0.0 160.2

176 74.5 0.0 155.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.12


300000


200000

100000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX018335.D
Acq: 10 Sep 2020 11:42

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01

Tot Ion:117 Resp: 646485
Ion Ratio Lower Upper
117 100
82 54.6 44.2 66.2
119 32.7 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX018335.D
Acq: 10 Sep 2020 11:42

Tgt Ion:152 Resp: 298627
Ion Ratio Lower Upper
152 100
115 58.9 41.5 124.5
150 154.4 0.0 343.8

