

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024167.D
 Acq On : 10 Sep 2021 15:45
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 11 04:24:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	150264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	245611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107353	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106931	50.987	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.980%
35) Dibromofluoromethane	5.397	113	84475	49.577	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.160%
50) Toluene-d8	8.653	98	312967	48.259	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.520%
62) 4-Bromofluorobenzene	11.085	95	118742	49.738	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.480%
Target Compounds						
16) Acetone	2.386	43	1039	1.108	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024167.D
Acq On : 10 Sep 2021 15:45
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 11 04:24:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

