

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024170.D
 Acq On : 10 Sep 2021 16:56
 Operator : JC/MD
 Sample : M3689-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 253

Quant Time: Sep 11 04:25:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107768	51.849	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.700%
35) Dibromofluoromethane	5.397	113	83168	49.928	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.860%
50) Toluene-d8	8.653	98	305879	48.246	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.085	95	115688	49.568	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.140%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1310	1.410	ug/l #	70
19) Methyl tert-butyl Ether	3.123	73	5529	1.112	ug/l	96
20) Methylene Chloride	2.794	84	2064	1.163	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024170.D
Acq On : 10 Sep 2021 16:56
Operator : JC/MD
Sample : M3689-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
253

Quant Time: Sep 11 04:25:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

