

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024172.D
 Acq On : 10 Sep 2021 17:43
 Operator : JC/MD
 Sample : PB139005TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139005TB

Quant Time: Sep 11 04:25:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100720	51.825	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.660%
35) Dibromofluoromethane	5.397	113	77957	49.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	8.653	98	292943	48.975	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.085	95	107828	48.969	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	52672	60.614	ug/l	92
20) Methylene Chloride	2.800	84	7031	4.236	ug/l	91
25) 2-Butanone	4.574	43	12529	9.519	ug/l #	86
43) Isopropyl Acetate	6.355	43	22181	5.246	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024172.D
Acq On : 10 Sep 2021 17:43
Operator : JC/MD
Sample : PB139005TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139005TB

Quant Time: Sep 11 04:25:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

