

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024173.D
 Acq On : 10 Sep 2021 18:06
 Operator : JC/MD
 Sample : M3686-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MAIN-GUARD-HOUSE-JET-GROUT

Quant Time: Sep 11 04:26:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	146239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103985	50.947	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.900%
35) Dibromofluoromethane	5.397	113	34465	21.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	42.540%#
50) Toluene-d8	8.653	98	303671	49.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.085	95	113078	49.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.640%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	161983	177.496	ug/l	88
20) Methylene Chloride	2.794	84	14414	8.269	ug/l	90
25) 2-Butanone	4.574	43	12318	8.911	ug/l	98
95) Naphthalene	13.780	128	9845	2.578	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024173.D
Acq On : 10 Sep 2021 18:06
Operator : JC/MD
Sample : M3686-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MAIN-GUARD-HOUSE-JET-GROUT

Quant Time: Sep 11 04:26:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

