

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024155.D
 Acq On : 10 Sep 2021 10:48
 Operator : JC/MD
 Sample : VX0910WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBL01

Quant Time: Sep 11 04:19:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	181204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	303916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122529	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	135809	53.700	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.400%
35) Dibromofluoromethane	5.397	113	107050	53.218	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.440%
50) Toluene-d8	8.653	98	389470	50.871	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.740%
62) 4-Bromofluorobenzene	11.085	95	146536	51.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.980%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024155.D
Acq On : 10 Sep 2021 10:48
Operator : JC/MD
Sample : VX0910WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01

Quant Time: Sep 11 04:19:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01


Tgt Ion:168 Resp: 181204
Ion Ratio Lower Upper
168 100
99 60.4 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 53.700 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48

Tgt Ion: 65 Resp: 135809
Ion Ratio Lower Upper
65 100
67 50.7 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01


Tgt Ion:114 Resp: 303916
Ion Ratio Lower Upper
114 100
63 20.9 0.0 40.8
88 17.2 0.0 32.8



#35
Dibromofluoromethane
Concen: 53.218 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48

Tgt Ion:113 Resp: 107050
Ion Ratio Lower Upper
113 100
111 102.0 81.9 122.9
192 19.6 16.7 25.1









#50
Toluene-d8
Concen: 50.871 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01


Tgt Ion: 98 Resp: 389470
Ion Ratio Lower Upper
98 100
100 64.0 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 51.993 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48

Tgt Ion: 95 Resp: 146536
Ion Ratio Lower Upper
95 100
174 78.9 0.0 154.6
176 76.5 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBL01

Tgt Ion:117 Resp: 287753
Ion Ratio Lower Upper
117 100
82 57.7 47.4 71.2
119 31.2 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024155.D
Acq: 10 Sep 2021 10:48

Tgt Ion:152 Resp: 122529
Ion Ratio Lower Upper
152 100
115 60.6 41.1 123.3
150 156.2 0.0 349.0

