

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024156.D
 Acq On : 10 Sep 2021 11:15
 Operator : JC/MD
 Sample : VX0910WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBS01

Manual Integrations
 APPROVED

MMDadoda
 9/13/2021 7:33:37 PM

Quant Time: Sep 11 04:19:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	181460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	279655	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127777	50.453	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.900%	
35) Dibromofluoromethane	5.391	113	97147	52.485	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.960%	
50) Toluene-d8	8.653	98	345132	48.990	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.980%	
62) 4-Bromofluorobenzene	11.085	95	142414	54.914	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.820%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	36722	19.181	ug/l	100
3) Chloromethane	1.294	50	34393	16.364	ug/l	99
4) Vinyl Chloride	1.374	62	33682	16.432	ug/l	99
5) Bromomethane	1.599	94	25072	19.455	ug/l	97
6) Chloroethane	1.685	64	22386	15.145	ug/l	100
7) Trichlorofluoromethane	1.886	101	59567	18.402	ug/l	94
8) Diethyl Ether	2.136	74	19977	17.672	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.331	101	34904	19.163	ug/l	97
10) Methyl Iodide	2.453	142	38419	15.148	ug/l #	89
11) Tert butyl alcohol	2.971	59	54619	106.642	ug/l	99
12) 1,1-Dichloroethene	2.319	96	32116	18.442	ug/l	94
13) Acrolein	2.239	56	25401	83.340	ug/l	97
14) Allyl chloride	2.666	41	54851	17.091	ug/l	89
15) Acrylonitrile	3.068	53	98669	86.463	ug/l	97
16) Acetone	2.386	43	101797	89.895	ug/l	91
17) Carbon Disulfide	2.514	76	70881	17.065	ug/l	98
18) Methyl Acetate	2.709	43	65967	23.335	ug/l	91
19) Methyl tert-butyl Ether	3.123	73	116553	19.230	ug/l	99
20) Methylene Chloride	2.794	84	37314	17.251	ug/l #	87
21) trans-1,2-Dichloroethene	3.099	96	35073	18.855	ug/l	97
22) Diisopropyl ether	3.770	45	114746	17.951	ug/l	96
23) Vinyl Acetate	3.727	43	482124	93.240	ug/l #	93
24) 1,1-Dichloroethane	3.617	63	64902	18.244	ug/l	99
25) 2-Butanone	4.568	43	149238	87.010	ug/l	95
26) 2,2-Dichloropropane	4.483	77	58988	20.203	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	41723	18.850	ug/l	97
28) Bromochloromethane	4.910	49	26760	16.544	ug/l	98
29) Tetrahydrofuran	5.019	42	94874	85.397	ug/l	91
30) Chloroform	5.105	83	70724	18.776	ug/l	95
31) Cyclohexane	5.477	56	57504	17.673	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	63757	19.767	ug/l	99
36) 1,1-Dichloropropene	5.696	75	51485	19.880	ug/l	98
37) Ethyl Acetate	4.727	43	55741	17.510	ug/l	95
38) Carbon Tetrachloride	5.684	117	53948	19.976	ug/l	98
39) Methylcyclohexane	7.385	83	62867	19.937	ug/l	95
40) Benzene	6.043	78	147204	19.222	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
 Data File : VX024156.D
 Acq On : 10 Sep 2021 11:15
 Operator : JC/MD
 Sample : VX0910WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBS01

Manual Integrations
 APPROVED

MMDadoda
 9/13/2021 7:33:37 PM

Quant Time: Sep 11 04:19:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	30826	17.862	ug/l	92
42) 1,2-Dichloroethane	6.092	62	62290	20.696	ug/l	94
43) Isopropyl Acetate	6.348	43	90127	18.100	ug/l	94
44) Trichloroethene	7.129	130	41097	19.427	ug/l	95
45) 1,2-Dichloropropane	7.433	63	39160	19.257	ug/l	94
46) Dibromomethane	7.586	93	27157	19.191	ug/l	99
47) Bromodichloromethane	7.824	83	50972	19.911	ug/l	98
48) Methyl methacrylate	7.696	41	54764	22.545	ug/l	87
49) 1,4-Dioxane	7.665	88	19530	380.414	ug/l #	95
51) 4-Methyl-2-Pentanone	8.580	43	296356	92.069	ug/l	90
52) Toluene	8.720	92	96337	19.414	ug/l	97
53) t-1,3-Dichloropropene	8.982	75	57786	18.865	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	59464	19.169	ug/l #	90
55) 1,1,2-Trichloroethane	9.153	97	40054	19.606	ug/l	98
56) Ethyl methacrylate	9.122	69	47747	14.230	ug/l #	85
57) 1,3-Dichloropropane	9.311	76	65437	19.312	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	156776	95.300	ug/l	99
59) 2-Hexanone	9.433	43	234668	94.844	ug/l	87
60) Dibromochloromethane	9.525	129	39779	18.723	ug/l	100
61) 1,2-Dibromoethane	9.610	107	42931	19.892	ug/l	99
64) Tetrachloroethene	9.275	164	42349	20.058	ug/l	95
65) Chlorobenzene	10.079	112	107465	19.482	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	39315	20.025	ug/l	99
67) Ethyl Benzene	10.195	91	194043	20.144	ug/l	99
68) m/p-Xylenes	10.305	106	144431	39.366	ug/l	98
69) o-Xylene	10.646	106	70412	19.529	ug/l	100
70) Styrene	10.659	104	104999	17.876	ug/l	96
71) Bromoform	10.805	173	28351	19.016	ug/l #	100
73) Isopropylbenzene	10.963	105	193071	19.163	ug/l	100
74) N-amyl acetate	10.848	43	78650	17.273	ug/l	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	62878	19.043	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	59517m	19.315	ug/l	
77) Bromobenzene	11.201	156	47391	19.097	ug/l	98
78) n-propylbenzene	11.305	91	217802	18.900	ug/l	100
79) 2-Chlorotoluene	11.366	91	132544	19.030	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	166293	20.084	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.024	75	18606	17.762	ug/l	88
82) 4-Chlorotoluene	11.457	91	154913	19.594	ug/l	99
83) tert-Butylbenzene	11.719	119	154446	19.206	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	168163	20.369	ug/l	97
85) sec-Butylbenzene	11.896	105	186540	18.322	ug/l	100
86) p-Isopropyltoluene	12.012	119	168686	20.131	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	84928	18.916	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	83893	18.735	ug/l	97
89) n-Butylbenzene	12.335	91	148700	19.875	ug/l	97
90) Hexachloroethane	12.542	117	25083	18.135	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	82952	19.071	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.945	75	12568	17.693	ug/l	89
93) 1,2,4-Trichlorobenzene	13.591	180	49959	19.084	ug/l	99
94) Hexachlorobutadiene	13.731	225	24068	20.628	ug/l	98
95) Naphthalene	13.780	128	161752	17.594	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	50856	19.506	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091021\
Data File : VX024156.D
Acq On : 10 Sep 2021 11:15
Operator : JC/MD
Sample : VX0910WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01


Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM

Quant Time: Sep 11 04:19:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration





Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 168 Resp: 181460
Ion Ratio Lower Upper
168 100
99 59.6 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 19.181 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.001 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 85 Resp: 36722
Ion Ratio Lower Upper
85 100
87 32.9 16.4 49.2




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED





MMDadoda
9/13/2021 7:33:37 PM

Abundance Scan 34 (1.294 min): VX023756.D\data.ms (-30)






#3
Chloromethane
Concen: 16.364 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 50 Resp: 34393
Ion Ratio Lower Upper
50 100
52 33.8 26.4 39.6



#4
Vinyl Chloride
Concen: 16.432 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion: 62 Resp: 33682
Ion Ratio Lower Upper
62 100
64 32.4 25.7 38.5



Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM







#5
Bromomethane
Concen: 19.455 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 94 Resp: 25072
Ion Ratio Lower Upper
94 100
96 92.3 76.4 114.6

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#6
Chloroethane
Concen: 15.145 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion: 64 Resp: 22386
Ion Ratio Lower Upper
64 100
66 31.6 25.4 38.0





#7
 Trichlorofluoromethane
 Concen: 18.402 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.000 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBS01




Tgt Ion:101 Resp: 59567
 Ion Ratio Lower Upper
 101 100
 103 62.0 53.2 79.8

Manual Integrations
 APPROVED



MMDadoda
 9/13/2021 7:33:37 PM




#8
 Diethyl Ether
 Concen: 17.672 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.006 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15



Tgt Ion: 74 Resp: 19977
 Ion Ratio Lower Upper
 74 100
 45 96.0 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.163 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	101	Resp:	34904
Ion	Ratio	Lower	Upper
101	100		
85	45.4	34.9	52.3
151	79.1	61.4	92.2

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#10
Methyl Iodide
Concen: 15.148 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion:	142	Resp:	38419
Ion	Ratio	Lower	Upper
142	100		
127	50.5	33.3	49.9#
141	14.4	11.4	17.0





#11
Tert butyl alcohol
Concen: 106.642 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.012 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Tgt Ion: 59 Resp: 54619
Ion Ratio Lower Upper
59 100
57 9.6 7.8 11.8

Manual Integrations
APPROVED



MMDadoda
9/13/2021 7:33:37 PM




#12
1,1-Dichloroethene
Concen: 18.442 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15



Tgt Ion: 96 Resp: 32116
Ion Ratio Lower Upper
96 100
61 166.2 124.6 187.0
98 61.8 50.6 76.0









#13
Acrolein
Concen: 83.340 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 56 Resp: 25401
Ion Ratio Lower Upper
56 100
55 72.6 56.4 84.6

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#14
Allyl chloride
Concen: 17.091 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 41 Resp: 54851
Ion Ratio Lower Upper
41 100
39 73.6 51.6 77.4
76 32.8 30.5 45.7









#15
Acrylonitrile
Concen: 86.463 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.007 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	53	Resp:	98669
Ion	Ratio	Lower	Upper
53	100		
52	84.2	65.3	97.9
51	37.7	28.2	42.2

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#16
Acetone
Concen: 89.895 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	43	Resp:	101797
Ion	Ratio	Lower	Upper
43	100		
58	29.7	27.8	41.8









#17
Carbon Disulfide
Concen: 17.065 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 76 Resp: 70881
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#18
Methyl Acetate
Concen: 23.335 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 43 Resp: 65967
Ion Ratio Lower Upper
43 100
74 22.9 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 19.230 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 73 Resp: 116553
Ion Ratio Lower Upper
73 100
57 22.1 17.1 25.7

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#20
Methylene Chloride
Concen: 17.251 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 84 Resp: 37314
Ion Ratio Lower Upper
84 100
49 130.5 88.6 132.8
51 41.2 27.0 40.6#
86 64.1 50.3 75.5





#21
trans-1,2-Dichloroethene
Concen: 18.855 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion: 96 Resp: 35073

Ion	Ratio	Lower	Upper
96	100		
61	132.4	107.7	161.5
98	57.9	50.4	75.6

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED




MMDadoda
9/13/2021 7:33:37 PM




#22
Diisopropyl ether
Concen: 17.951 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 45 Resp: 114746

Ion	Ratio	Lower	Upper
45	100		
43	75.3	58.0	87.0
87	28.1	25.4	38.0
59	11.4	10.2	15.2









#23
 Vinyl Acetate
 Concen: 93.240 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.006 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBS01

Tgt Ion: 43 Resp: 482124
 Ion Ratio Lower Upper
 43 100
 86 9.7 9.9 14.9#




Manual Integrations
 APPROVED


MMDadoda
 9/13/2021 7:33:37 PM



#24
 1,1-Dichloroethane
 Concen: 18.244 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15

Tgt Ion: 63 Resp: 64902
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.8 11.2
 100 4.8 2.5 7.5








#25
2-Butanone
Concen: 87.010 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED




MMDadoda
9/13/2021 7:33:37 PM


Tgt Ion: 43 Resp: 149238
Ion Ratio Lower Upper
43 100
72 26.8 23.7 35.5



#26
2,2-Dichloropropane
Concen: 20.203 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 77 Resp: 58988
Ion Ratio Lower Upper
77 100
97 21.8 11.7 35.1





#27

cis-1,2-Dichloroethene
Concen: 18.850 ug/l
RT: 4.495 min Scan# 559
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion: 96 Resp: 41723

Ion	Ratio	Lower	Upper
96	100		
61	147.3	0.0	285.6
98	63.5	0.0	129.6

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM







#28

Bromochloromethane
Concen: 16.544 ug/l
RT: 4.910 min Scan# 627
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 49 Resp: 26760


Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.0
130	83.8	68.2	102.2





#29
Tetrahydrofuran
Concen: 85.397 ug/l
RT: 5.019 min Scan# 645
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Tgt Ion: 42 Resp: 94874
Ion Ratio Lower Upper
42 100
72 45.3 41.7 62.5
71 42.9 38.4 57.6

Manual Integrations
APPROVED



MMDadoda
9/13/2021 7:33:37 PM




#30
Chloroform
Concen: 18.776 ug/l
RT: 5.105 min Scan# 659
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Tgt Ion: 83 Resp: 70724
Ion Ratio Lower Upper
83 100
85 61.1 52.0 78.0









#31
Cyclohexane
Concen: 17.673 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion: 56 Resp: 57504
Ion Ratio Lower Upper
56 100
69 33.2 25.3 37.9
84 90.2 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 19.767 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Tgt Ion: 97 Resp: 63757
Ion Ratio Lower Upper
97 100
99 64.5 51.5 77.3
61 44.2 34.2 51.4



Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#33





1,2-Dichloroethane-d4
Concen: 50.453 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 65 Resp: 127777
Ion Ratio Lower Upper
65 100
67 49.9 0.0 105.4

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED




MMDadoda
9/13/2021 7:33:37 PM




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 114 Resp: 279655
Ion Ratio Lower Upper
114 100
63 20.6 0.0 40.8
88 16.1 0.0 32.8








#35
Dibromofluoromethane
Concen: 52.485 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:113 Resp: 97147
Ion Ratio Lower Upper
113 100
111 103.2 81.9 122.9
192 19.3 16.7 25.1

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#36
1,1-Dichloropropene
Concen: 19.880 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 75 Resp: 51485
Ion Ratio Lower Upper
75 100
110 35.7 18.1 54.4
77 30.0 25.3 37.9









#37
Ethyl Acetate
Concen: 17.510 ug/l
RT: 4.727 min Scan# 597
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	43	Resp:	55741
Ion Ratio	Lower	Upper	
43	100		
61	13.7	12.8	19.2
70	11.1	10.5	15.7

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#38
Carbon Tetrachloride
Concen: 19.976 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	117	Resp:	53948
Ion Ratio	Lower	Upper	
117	100		
119	94.0	76.2	114.2
121	29.1	24.8	37.2





#39

Methylcyclohexane

Concen: 19.937 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX024156.D





Acq: 10 Sep 2021 11:15

Tgt Ion:	83	Resp:	62867
Ion	Ratio	Lower	Upper
83	100		
55	68.8	58.9	88.3
98	45.0	38.1	57.1

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#40

Benzene

Concen: 19.222 ug/l




RT: 6.043 min Scan# 813

Delta R.T. -0.007 min

Lab File: VX024156.D


Acq: 10 Sep 2021 11:15

Tgt Ion:	78	Resp:	147204
Ion	Ratio	Lower	Upper
78	100		
77	23.9	18.9	28.3









#41
Methacrylonitrile
Concen: 17.862 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion:	41	Resp:	30826
Ion	Ratio	Lower	Upper
41	100		
39	58.6	42.5	63.7
67	72.1	65.0	97.6
52	31.4	25.3	37.9



#42
1,2-Dichloroethane
Concen: 20.696 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Tgt Ion:	62	Resp:	62290
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	20.2



Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM







#43
Isopropyl Acetate
Concen: 18.100 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 43 Resp: 90127
Ion Ratio Lower Upper
43 100
61 19.7 18.2 27.4
87 13.4 12.5 18.7

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#44
Trichloroethene
Concen: 19.427 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:130 Resp: 41097
Ion Ratio Lower Upper
130 100
95 94.3 0.0 198.0





#45





1,2-Dichloropropane
Concen: 19.257 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. -0.007 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 63 Resp: 39160
Ion Ratio Lower Upper
63 100
65 29.0 25.7 38.5

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED




MMDadoda
9/13/2021 7:33:37 PM



#46

Dibromomethane
Concen: 19.191 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion: 93 Resp: 27157
Ion Ratio Lower Upper
93 100
95 84.4 66.8 100.2
174 102.2 83.1 124.7





#47
Bromodichloromethane
Concen: 19.911 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01



Tgt Ion: 83 Resp: 50972
Ion Ratio Lower Upper
83 100
85 65.7 51.4 77.0
127 7.9 6.8 10.2




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#48
Methyl methacrylate
Concen: 22.545 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 41 Resp: 54764
Ion Ratio Lower Upper
41 100
69 79.3 76.0 114.0
39 61.3 44.6 67.0









#49
1,4-Dioxane
Concen: 380.414 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 88 Resp: 19530
Ion Ratio Lower Upper
88 100
43 35.9 23.2 34.8#
58 69.6 55.0 82.4

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#50
Toluene-d8
Concen: 48.990 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 98 Resp: 345132
Ion Ratio Lower Upper
98 100
100 62.7 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 92.069 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 43 Resp: 296356
Ion Ratio Lower Upper
43 100
58 37.9 35.4 53.0

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#52
Toluene
Concen: 19.414 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 92 Resp: 96337
Ion Ratio Lower Upper
92 100
91 175.0 137.1 205.7








#53
t-1,3-Dichloropropene
Concen: 18.865 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 75 Resp: 57786
Ion Ratio Lower Upper
75 100
77 32.1 24.9 37.3

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#54
cis-1,3-Dichloropropene
Concen: 19.169 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 75 Resp: 59464
Ion Ratio Lower Upper
75 100
77 31.9 25.7 38.5
39 54.8 35.1 52.7#





#55





1,1,2-Trichloroethane
Concen: 19.606 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	97	Resp:	40054
Ion	Ratio	Lower	Upper
97	100		
83	84.6	69.3	103.9
85	55.3	44.3	66.5
99	60.2	50.3	75.5

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED




MMDadoda
9/13/2021 7:33:37 PM




#56

Ethyl methacrylate
Concen: 14.230 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	69	Resp:	47747
Ion	Ratio	Lower	Upper
69	100		
41	67.5	46.8	70.2
39	42.6	24.7	37.1#









#57
1,3-Dichloropropane
Concen: 19.312 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 76 Resp: 65437
Ion Ratio Lower Upper
76 100
78 33.0 26.2 39.2

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#58
2-Chloroethyl Vinyl ether
Concen: 95.300 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 63 Resp: 156776
Ion Ratio Lower Upper
63 100
106 29.0 23.5 35.3









#59
2-Hexanone
Concen: 94.844 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 43 Resp: 234668
Ion Ratio Lower Upper
43 100
58 51.9 30.9 92.8

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#60
Dibromochloromethane
Concen: 18.723 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion: 129 Resp: 39779
Ion Ratio Lower Upper
129 100
127 76.9 38.4 115.2





#61
1,2-Dibromoethane
Concen: 19.892 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01



Tgt Ion:107 Resp: 42931
Ion Ratio Lower Upper
107 100
109 93.5 75.8 113.6




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#62
4-Bromofluorobenzene
Concen: 54.914 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 95 Resp: 142414
Ion Ratio Lower Upper
95 100
174 78.9 0.0 154.6
176 77.5 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Tgt Ion:117 Resp: 264342
Ion Ratio Lower Upper
117 100
82 57.0 47.4 71.2
119 31.2 26.6 39.8




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#64
Tetrachloroethene
Concen: 20.058 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:164 Resp: 42349
Ion Ratio Lower Upper
164 100
166 123.0 103.4 155.2
129 96.3 70.8 106.2
131 90.7 72.5 108.7









#65
Chlorobenzene
Concen: 19.482 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Tgt Ion:112 Resp: 107465
Ion Ratio Lower Upper
112 100
114 31.7 25.8 38.6




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 20.025 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:131 Resp: 39315
Ion Ratio Lower Upper
131 100
133 94.7 47.7 143.1
119 64.8 32.1 96.5









#67
Ethyl Benzene
Concen: 20.144 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 91 Resp: 194043
Ion Ratio Lower Upper
91 100
106 31.2 24.6 37.0

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#68
m/p-Xylenes
Concen: 39.366 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 106 Resp: 144431
Ion Ratio Lower Upper
106 100
91 206.4 162.8 244.2









#69
o-Xylene
Concen: 19.529 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Tgt Ion:106 Resp: 70412
Ion Ratio Lower Upper
106 100
91 216.0 108.1 324.2




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM




#70
Styrene
Concen: 17.876 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:104 Resp: 104999
Ion Ratio Lower Upper
104 100
78 54.9 40.0 60.0
103 55.4 43.4 65.0









#71
Bromoform
Concen: 19.016 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion:173 Resp: 28351
Ion Ratio Lower Upper
173 100
175 47.8 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15



Tgt Ion:152 Resp: 132093
Ion Ratio Lower Upper
152 100
115 71.5 41.1 123.3
150 162.2 0.0 349.0



Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM







#73
Isopropylbenzene
Concen: 19.163 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Tgt Ion:105 Resp: 193071
Ion Ratio Lower Upper
105 100
120 26.2 13.2 39.6




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#74
N-ethyl acetate
Concen: 17.273 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 43 Resp: 78650
Ion Ratio Lower Upper
43 100
70 41.5 39.0 58.6
55 24.7 21.5 32.3
61 22.4 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 19.043 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 83 Resp: 62878
Ion Ratio Lower Upper
83 100
131 10.8 5.1 15.2
85 64.0 31.8 95.4

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM




#76
1,2,3-Trichloropropane
Concen: 19.315 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 75 Resp: 59517
Ion Ratio Lower Upper
75 100
77 40.4 19.1 57.5









#77
Bromobenzene
Concen: 19.097 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15





Tgt Ion: 156 Resp: 47391
Ion Ratio Lower Upper
156 100
77 161.4 82.8 248.6
158 99.3 49.1 147.4



#78
n-propylbenzene
Concen: 18.900 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15



Tgt Ion: 91 Resp: 217802
Ion Ratio Lower Upper
91 100
120 23.0 11.5 34.5



Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM







#79
2-Chlorotoluene
Concen: 19.030 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 91 Resp: 132544
Ion Ratio Lower Upper
91 100
126 33.5 16.6 49.8

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#80
1,3,5-Trimethylbenzene
Concen: 20.084 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion:105 Resp: 166293
Ion Ratio Lower Upper
105 100
120 46.5 23.8 71.5





#81
trans-1,4-Dichloro-2-butene
Concen: 17.762 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampled :
VX0910WBS01




Tgt Ion: 75 Resp: 18606
Ion Ratio Lower Upper
75 100
53 108.7 74.5 111.7
89 45.7 37.9 56.9

Manual Integrations
APPROVED



MMDadoda
9/13/2021 7:33:37 PM




#82
4-Chlorotoluene
Concen: 19.594 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Tgt Ion: 91 Resp: 154913
Ion Ratio Lower Upper
91 100
126 29.0 14.9 44.5





#83
tert-Butylbenzene
Concen: 19.206 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Tgt Ion:119 Resp: 154446
Ion Ratio Lower Upper
119 100
91 62.1 27.9 83.7
134 23.6 11.3 33.9

Manual Integrations
APPROVED



MMDadoda
9/13/2021 7:33:37 PM




#84
1,2,4-Trimethylbenzene
Concen: 20.369 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15



Tgt Ion:105 Resp: 168163
Ion Ratio Lower Upper
105 100
120 43.2 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 18.322 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0910WBS01

Tgt Ion:105 Resp: 186540
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.0 29.8




Manual Integrations
 APPROVED


MMDadoda
 9/13/2021 7:33:37 PM



#86
 p-Isopropyltoluene
 Concen: 20.131 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX024156.D
 Acq: 10 Sep 2021 11:15

Tgt Ion:119 Resp: 168686
 Ion Ratio Lower Upper
 119 100
 134 26.3 12.9 38.7
 91 25.7 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 18.916 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:146 Resp: 84928
Ion Ratio Lower Upper
146 100
111 41.5 20.2 60.5
148 64.5 31.9 95.7

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#88
1,4-Dichlorobenzene
Concen: 18.735 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.001 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:146 Resp: 83893
Ion Ratio Lower Upper
146 100
111 42.2 19.6 58.8
148 64.5 31.6 95.0








#89
n-Butylbenzene
Concen: 19.875 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01

Tgt Ion: 91 Resp: 148700
Ion Ratio Lower Upper
91 100
92 51.7 27.4 82.0
134 25.4 13.2 39.5




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#90
Hexachloroethane
Concen: 18.135 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion: 117 Resp: 25083
Ion Ratio Lower Upper
117 100
201 80.2 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 19.071 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15

Tgt Ion:	146	Resp:	82952
Ion Ratio	Lower	Upper	
146	100		
111	42.0	20.8	62.2
148	63.6	31.9	95.5

Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Manual Integrations
APPROVED


MMDadoda
9/13/2021 7:33:37 PM



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.693 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15


Tgt Ion:	75	Resp:	12568
Ion Ratio	Lower	Upper	
75	100		
155	79.7	45.3	135.9
157	103.1	57.7	173.1





#93
1,2,4-Trichlorobenzene
Concen: 19.084 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Tgt Ion:180 Resp: 49959
Ion Ratio Lower Upper
180 100
182 95.4 47.7 143.1
145 32.5 15.0 45.0

Manual Integrations
APPROVED



MMDadoda
9/13/2021 7:33:37 PM




#94
Hexachlorobutadiene
Concen: 20.628 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Tgt Ion:225 Resp: 24068
Ion Ratio Lower Upper
225 100
223 63.5 31.1 93.2
227 65.4 31.6 95.0





#95
Naphthalene
Concen: 17.594 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15




Instrument :
MSVOA_X
ClientSampleId :
VX0910WBS01




Tgt Ion:128 Resp: 161752
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 10.6 8.8 13.2

Manual Integrations
APPROVED

MMDadoda
9/13/2021 7:33:37 PM



#96
1,2,3-Trichlorobenzene
Concen: 19.506 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX024156.D
Acq: 10 Sep 2021 11:15



Tgt Ion:180 Resp: 50856
Ion Ratio Lower Upper
180 100
182 94.3 47.6 142.9
145 33.3 16.1 48.3

