

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091118\
 Data File : VX004476.D
 Acq On : 11 Sep 2018 20:07
 Operator : JC/MD
 Sample : J4857-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS033-180910

Quant Time: Sep 12 11:25:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 15:21:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	196375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	313683	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304080	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167230	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141288	55.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.60%	
35) Dibromofluoromethane	5.51	113	124188	43.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.52%	
50) Toluene-d8	8.72	98	454036	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
62) 4-Bromofluorobenzene	11.14	95	159970	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
16) Acetone	2.45	43	30796	24.92	ug/l	Qvalue 100
27) cis-1,2-Dichloroethene	4.60	96	7367	2.91	ug/l	94
44) Trichloroethene	7.22	130	1156	0.39	ug/l	76
64) Tetrachloroethene	9.34	164	2334	0.61	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091118\
Data File : VX004476.D
Acq On : 11 Sep 2018 20:07
Operator : JC/MD
Sample : J4857-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS033-180910

Quant Time: Sep 12 11:25:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 15:21:13 2018
Response via : Initial Calibration

