

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091118\
 Data File : VX004479.D
 Acq On : 11 Sep 2018 21:28
 Operator : JC/MD
 Sample : J4842-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SG-SOIL

Quant Time: Sep 12 07:42:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 15:21:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	186746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	305606	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	300217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	161986	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	138404	57.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.96%	
35) Dibromofluoromethane	5.51	113	119361	42.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.36%	
50) Toluene-d8	8.72	98	444093	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
62) 4-Bromofluorobenzene	11.14	95	154582	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
Target Compounds						
16) Acetone	2.45	43	12267	10.44	ug/l	97
18) Methyl Acetate	2.78	43	5164	1.57	ug/l	98
20) Methylene Chloride	2.85	84	376771	156.28	ug/l	96
25) 2-Butanone	4.71	43	2966	1.59	ug/l	93
43) Isopropyl Acetate	6.46	43	120156	21.44	ug/l	98
95) Naphthalene	13.83	128	19347	1.75	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091118\
Data File : VX004479.D
Acq On : 11 Sep 2018 21:28
Operator : JC/MD
Sample : J4842-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SG-SOIL

Quant Time: Sep 12 07:42:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 15:21:13 2018
Response via : Initial Calibration

