

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091119\
 Data File : VX012301.D
 Acq On : 11 Sep 2019 14:41
 Operator : SY/SP
 Sample : K4643-09
 Misc : 3.09g/5.0mL/MSVOA X/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-B4-(4)

Quant Time: Sep 12 06:31:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 12:55:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	284136	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	111886	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
35) Dibromofluoromethane	5.49	113	84585	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
50) Toluene-d8	8.71	98	332426	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	11.14	95	123872	41.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.64%	
Target Compounds						
16) Acetone	2.43	43	59064	107.795	ug/l	98
25) 2-Butanone	4.67	43	8481	12.108	ug/l	100
29) Tetrahydrofuran	5.13	42	4405	11.656	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091119\
Data File : VX012301.D
Acq On : 11 Sep 2019 14:41
Operator : SY/SP
Sample : K4643-09
Misc : 3.09g/5.0mL/MSVOA X/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-B4-(4)

Quant Time: Sep 12 06:31:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091019S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 12:55:26 2019
Response via : Initial Calibration

